

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN090122\
 Data File : VN074245.D
 Acq On : 01 Sep 2022 16:19
 Operator : JC\MD
 Sample : N4329-01
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 MW-3R

Quant Time: Sep 02 05:39:34 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N083022W.M
 Quant Title : SW846 8260
 QLast Update : Wed Aug 31 07:02:40 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.022	168	252602	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.904	114	467027	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.686	117	452425	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.610	152	195894	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.369	65	190681	48.662	ug/l	0.00
Spiked Amount 50.000	Range 74 - 125		Recovery =	97.320%		
35) Dibromofluoromethane	7.951	113	154325	50.781	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery =	101.560%		
50) Toluene-d8	10.380	98	561206	46.892	ug/l	0.00
Spiked Amount 50.000	Range 86 - 113		Recovery =	93.780%		
62) 4-Bromofluorobenzene	12.668	95	214851	59.032	ug/l	0.00
Spiked Amount 50.000	Range 83 - 123		Recovery =	118.060%		
Target Compounds					Qvalue	
11) Tert butyl alcohol	5.304	59	31842	29.217	ug/l #	72
16) Acetone	4.222	43	8240	3.147	ug/l #	83
19) Methyl tert-butyl Ether	5.534	73	1277410	116.606	ug/l	93
22) Diisopropyl ether	6.410	45	62745	4.663	ug/l #	89
27) cis-1,2-Dichloroethene	7.251	96	37479	9.816	ug/l	95
31) Cyclohexane	8.016	56	17254	2.665	ug/l #	84
39) Methylcyclohexane	9.404	83	6277	1.084	ug/l #	79
40) Benzene	8.392	78	1396368	90.190	ug/l	100
44) Trichloroethene	9.157	130	5175	1.496	ug/l	76
52) Toluene	10.439	92	2128	0.233	ug/l	84
64) Tetrachloroethene	10.910	164	1592	0.486	ug/l #	61
67) Ethyl Benzene	11.786	91	22623	1.238	ug/l #	82
68) m/p-Xylenes	11.892	106	3279	0.483	ug/l	85
73) Isopropylbenzene	12.521	105	8355	0.498	ug/l	92

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN090122\
Data File : VN074245.D
Acq On : 01 Sep 2022 16:19
Operator : JC\MD
Sample : N4329-01
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 18 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
MW-3R

Quant Time: Sep 02 05:39:34 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N083022W.M
Quant Title : SW846 8260
QLast Update : Wed Aug 31 07:02:40 2022
Response via : Initial Calibration

