

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN090122\
 Data File : VN074247.D
 Acq On : 01 Sep 2022 17:08
 Operator : JC\MD
 Sample : N4329-03
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 MW-14D

Quant Time: Sep 02 05:39:59 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N083022W.M
 Quant Title : SW846 8260
 QLast Update : Wed Aug 31 07:02:40 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.016	168	231423	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.904	114	425255	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.686	117	409321	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.610	152	185314	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.369	65	173528	48.337	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	96.680%
35) Dibromofluoromethane	7.957	113	142371	51.449	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	102.900%
50) Toluene-d8	10.380	98	514118	47.177	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	94.360%
62) 4-Bromofluorobenzene	12.669	95	189967	57.322	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	114.640%
Target Compounds						
					Qvalue	
16) Acetone	4.234	43	9587	3.997	ug/l	97
17) Carbon Disulfide	4.463	76	6731	0.872	ug/l #	94
27) cis-1,2-Dichloroethene	7.245	96	2592	0.741	ug/l	84

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN090122\
Data File : VN074247.D
Acq On : 01 Sep 2022 17:08
Operator : JC\MD
Sample : N4329-03
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 20 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
MW-14D

Quant Time: Sep 02 05:39:59 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N083022W.M
Quant Title : SW846 8260
QLast Update : Wed Aug 31 07:02:40 2022
Response via : Initial Calibration

