

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN090122\
 Data File : VN074249.D
 Acq On : 01 Sep 2022 17:57
 Operator : JC\MD
 Sample : N4329-05
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 MW-34

Quant Time: Sep 02 05:40:18 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N083022W.M
 Quant Title : SW846 8260
 QLast Update : Wed Aug 31 07:02:40 2022
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.016	168	257120	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	8.904	114	474406	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.681	117	465770	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.616	152	213231	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.369	65	204008	51.148	ug/l	0.00
Spiked Amount	50.000	Range 74 - 125	Recovery	= 102.300%		
35) Dibromofluoromethane	7.957	113	163998	53.124	ug/l	0.00
Spiked Amount	50.000	Range 75 - 124	Recovery	= 106.240%		
50) Toluene-d8	10.381	98	578714	47.602	ug/l	0.00
Spiked Amount	50.000	Range 86 - 113	Recovery	= 95.200%		
62) 4-Bromofluorobenzene	12.669	95	228543	61.817	ug/l	0.00
Spiked Amount	50.000	Range 83 - 123	Recovery	= 123.640%#		
Target Compounds						
						Qvalue
3) Chloromethane	2.269	50	1211	0.323	ug/l	90
16) Acetone	4.228	43	19515	7.322	ug/l	90
19) Methyl tert-butyl Ether	5.534	73	33157	2.973	ug/l #	65
25) 2-Butanone	7.269	43	4976	1.167	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN090122\
Data File : VN074249.D
Acq On : 01 Sep 2022 17:57
Operator : JC\MD
Sample : N4329-05
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 22 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
MW-34

Quant Time: Sep 02 05:40:18 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N083022W.M
Quant Title : SW846 8260
QLast Update : Wed Aug 31 07:02:40 2022
Response via : Initial Calibration

