

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN090122\
 Data File : VN074269.D
 Acq On : 02 Sep 2022 02:55
 Operator : JC\MD
 Sample : N4354-02
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 44 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 MW-41

Quant Time: Sep 02 05:44:54 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N083022W.M
 Quant Title : SW846 8260
 QLast Update : Wed Aug 31 07:02:40 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.016	168	217199	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.904	114	405892	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.686	117	394246	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.616	152	183938	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.369	65	163604	48.557	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	97.120%
35) Dibromofluoromethane	7.951	113	135480	51.294	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	102.580%
50) Toluene-d8	10.381	98	479775	46.126	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	92.260%
62) 4-Bromofluorobenzene	12.669	95	185877	58.763	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	117.520%
Target Compounds						
					Qvalue	
16) Acetone	4.240	43	9845	4.373	ug/l #	76
52) Toluene	10.439	92	3790	0.478	ug/l #	74
67) Ethyl Benzene	11.786	91	7972	0.501	ug/l #	84
68) m/p-Xylenes	11.892	106	6746	1.141	ug/l	99
69) o-Xylene	12.216	106	3241	0.550	ug/l	85
84) 1,2,4-Trimethylbenzene	13.304	105	6113	0.494	ug/l	88
95) Naphthalene	15.433	128	56663	5.117	ug/l	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN090122\
Data File : VN074269.D
Acq On : 02 Sep 2022 02:55
Operator : JC\MD
Sample : N4354-02
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 44 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
MW-41

Quant Time: Sep 02 05:44:54 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N083022W.M
Quant Title : SW846 8260
QLast Update : Wed Aug 31 07:02:40 2022
Response via : Initial Calibration

