

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN090219\
 Data File : VN057643.D
 Acq On : 1 Sep 2019 14:47
 Operator : JC/SP
 Sample : K4527-35
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 T9-12-13-O-(0)

Quant Time: Sep 02 02:29:05 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N090119W.M
 Quant Title : SW846 8260
 QLast Update : Sat Aug 31 15:54:32 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.65	168	180396	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.58	114	308600	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	271431	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.35	152	94231	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.01	65	222477	54.59	ug/l	0.00
Spiked Amount 50.000			Recovery	=	109.18%	
35) Dibromofluoromethane	7.58	113	136379	50.61	ug/l	0.00
Spiked Amount 50.000			Recovery	=	101.22%	
50) Toluene-d8	10.08	98	539043	52.10	ug/l	0.00
Spiked Amount 50.000			Recovery	=	104.20%	
62) 4-Bromofluorobenzene	12.40	95	162525	41.31	ug/l	0.00
Spiked Amount 50.000			Recovery	=	82.62%	
Target Compounds						
16) Acetone	3.81	43	12258	10.733	ug/l	93
18) Methyl Acetate	4.31	43	7720	1.969	ug/l #	75
20) Methylene Chloride	4.52	84	6007	2.125	ug/l	95
43) Isopropyl Acetate	8.16	43	56373	8.040	ug/l #	94
95) Naphthalene	15.13	128	5336	8.040	ug/l #	92

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN090219\
Data File : VN057643.D
Acq On : 1 Sep 2019 14:47
Operator : JC/SP
Sample : K4527-35
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
T9-12-13-O-(0)

Quant Time: Sep 02 02:29:05 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N090119W.M
Quant Title : SW846 8260
QLast Update : Sat Aug 31 15:54:32 2019
Response via : Initial Calibration

