

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN090219\
 Data File : VN057655.D
 Acq On : 1 Sep 2019 19:50
 Operator : JC/SP
 Sample : K4516-05
 Misc : 7.57µ/5mL/100µL/5.00mL/MSVOA_N/MEOH
 ALS Vial : 24 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 T1-S-B4-(0)

Quant Time: Sep 02 03:21:17 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N090119W.M
 Quant Title : SW846 8260
 QLast Update : Sat Aug 31 15:54:32 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.65	168	174698	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.58	114	300258	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	249787	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.35	152	103921	50.00	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.01	65	213891	54.20	ug/l	0.00
Spiked Amount	50.000		Recovery	=	108.40%	
35) Dibromofluoromethane	7.58	113	121165	46.21	ug/l	0.00
Spiked Amount	50.000		Recovery	=	92.42%	
50) Toluene-d8	10.08	98	488796	48.55	ug/l	0.00
Spiked Amount	50.000		Recovery	=	97.10%	
62) 4-Bromofluorobenzene	12.40	95	152840	39.93	ug/l	0.00
Spiked Amount	50.000		Recovery	=	79.86%	

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

