

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN090220\
 Data File : VN063186.D
 Acq On : 2 Sep 2020 11:29
 Operator : JC/MD
 Sample : PB131350TB
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB131350TB

Quant Time: Sep 03 01:29:07 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N083120W.M
 Quant Title : SW846 8260
 QLast Update : Tue Sep 01 04:21:28 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.63	168	350262	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.55	114	682751	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.38	117	690784	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.32	152	268829	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.99	65	280203	48.94	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.88%	
35) Dibromofluoromethane	7.55	113	218873	49.97	ug/l	0.00
Spiked Amount 50.000			Recovery	=	99.94%	
50) Toluene-d8	10.06	98	873153	51.79	ug/l	0.00
Spiked Amount 50.000			Recovery	=	103.58%	
62) 4-Bromofluorobenzene	12.38	95	308386	52.29	ug/l	0.00
Spiked Amount 50.000			Recovery	=	104.58%	
Target Compounds						
16) Acetone	3.78	43	24737	15.198	ug/l	96
18) Methyl Acetate	4.29	43	5596	1.341	ug/l #	74
20) Methylene Chloride	4.51	84	12794	2.326	ug/l	90
43) Isopropyl Acetate	8.13	43	126870	11.852	ug/l #	88

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN090220\
Data File : VN063186.D
Acq On : 2 Sep 2020 11:29
Operator : JC/MD
Sample : PB131350TB
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 6 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB131350TB

Quant Time: Sep 03 01:29:07 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N083120W.M
Quant Title : SW846 8260
QLast Update : Tue Sep 01 04:21:28 2020
Response via : Initial Calibration

