

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN090221\
 Data File : VN068391.D
 Acq On : 2 Sep 2021 17:37
 Operator : JC/MD
 Sample : M3606-02
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 BPU-2

Quant Time: Sep 03 01:25:31 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N083021W.M
 Quant Title : SW846 8260
 QLast Update : Thu Sep 02 14:59:33 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.088	168	422785	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	8.971	114	703945	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.746	117	692833	50.000	ug/l	# 0.00
72) 1,4-Dichlorobenzene-d4	13.678	152	282670	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.440	65	241482	55.959	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	111.920%
35) Dibromofluoromethane	8.024	113	222286	49.804	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	99.600%
50) Toluene-d8	10.443	98	857977	51.352	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	102.700%
62) 4-Bromofluorobenzene	12.733	95	285969	52.155	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	104.320%
Target Compounds						
95) Naphthalene	15.507	128	6069	2.841	ug/l	# 92

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN090221\
Data File : VN068391.D
Acq On : 2 Sep 2021 17:37
Operator : JC/MD
Sample : M3606-02
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
BPU-2

Quant Time: Sep 03 01:25:31 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N083021W.M
Quant Title : SW846 8260
QLast Update : Thu Sep 02 14:59:33 2021
Response via : Initial Calibration

