

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_N\Data\VN090221\  
 Data File : VN068389.D  
 Acq On : 2 Sep 2021 16:52  
 Operator : JC/MD  
 Sample : VN0902WBS01  
 Misc : 5.00mL/MSVOA\_N/WATER  
 ALS Vial : 5 Sample Multiplier: 1

Instrument :  
 MSVOA\_N  
 ClientSampleId :  
 VN0902WBS01

Manual Integrations  
 APPROVED

MMDadoda  
 9/3/2021 5:23:55 PM

Quant Time: Sep 03 01:25:02 2021  
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_N\methods\82N083021W.M  
 Quant Title : SW846 8260  
 QLast Update : Thu Sep 02 14:59:33 2021  
 Response via : Initial Calibration

| Compound                     | R.T.           | QIon | Response   | Conc    | Units  | Dev(Min) |
|------------------------------|----------------|------|------------|---------|--------|----------|
| -----                        |                |      |            |         |        |          |
| Internal Standards           |                |      |            |         |        |          |
| 1) Pentafluorobenzene        | 8.088          | 168  | 649735     | 50.000  | ug/l   | # 0.00   |
| 34) 1,4-Difluorobenzene      | 8.968          | 114  | 964195     | 50.000  | ug/l   | 0.00     |
| 63) Chlorobenzene-d5         | 11.746         | 117  | 860141     | 50.000  | ug/l   | # 0.00   |
| 72) 1,4-Dichlorobenzene-d4   | 13.678         | 152  | 423995     | 50.000  | ug/l   | 0.00     |
|                              |                |      |            |         |        |          |
| System Monitoring Compounds  |                |      |            |         |        |          |
| 33) 1,2-Dichloroethane-d4    | 8.440          | 65   | 316556     | 47.733  | ug/l   | 0.00     |
| Spiked Amount 50.000         | Range 61 - 141 |      | Recovery = | 95.460% |        |          |
| 35) Dibromofluoromethane     | 8.024          | 113  | 289205     | 47.308  | ug/l   | 0.00     |
| Spiked Amount 50.000         | Range 69 - 133 |      | Recovery = | 94.620% |        |          |
| 50) Toluene-d8               | 10.443         | 98   | 1066922    | 46.622  | ug/l   | 0.00     |
| Spiked Amount 50.000         | Range 65 - 126 |      | Recovery = | 93.240% |        |          |
| 62) 4-Bromofluorobenzene     | 12.733         | 95   | 336011     | 44.741  | ug/l   | 0.00     |
| Spiked Amount 50.000         | Range 58 - 135 |      | Recovery = | 89.480% |        |          |
|                              |                |      |            |         |        |          |
| Target Compounds             |                |      |            |         |        | Qvalue   |
| 2) Dichlorodifluoromethane   | 2.075          | 85   | 107411     | 19.865  | ug/l   | 99       |
| 3) Chloromethane             | 2.309          | 50   | 100705     | 17.772  | ug/l   | 98       |
| 4) Vinyl Chloride            | 2.445          | 62   | 161181     | 18.655  | ug/l   | 97       |
| 5) Bromomethane              | 2.845          | 94   | 146237     | 18.165  | ug/l   | 97       |
| 6) Chloroethane              | 3.014          | 64   | 118236     | 19.091  | ug/l   | 97       |
| 7) Trichlorofluoromethane    | 3.376          | 101  | 205650     | 11.978  | ug/l   | 97       |
| 8) Diethyl Ether             | 3.829          | 74   | 65081      | 13.160  | ug/l   | 81       |
| 9) 1,1,2-Trichlorotrifluo... | 4.224          | 101  | 116417m    | 21.485  | ug/l   |          |
| 10) Methyl Iodide            | 4.427          | 142  | 162735     | 23.537  | ug/l   | 89       |
| 11) Tert butyl alcohol       | 5.369          | 59   | 84882      | 97.881  | ug/l   | 99       |
| 12) 1,1-Dichloroethene       | 4.186          | 96   | 106998     | 21.120  | ug/l # | 76       |
| 13) Acrolein                 | 4.047          | 56   | 49383      | 87.679  | ug/l   | 96       |
| 14) Allyl chloride           | 4.846          | 41   | 139229     | 21.847  | ug/l # | 91       |
| 15) Acrylonitrile            | 5.554          | 53   | 230989     | 97.422  | ug/l   | 100      |
| 16) Acetone                  | 4.288          | 43   | 238937     | 129.177 | ug/l # | 84       |
| 17) Carbon Disulfide         | 4.537          | 76   | 269020     | 19.693  | ug/l   | 99       |
| 18) Methyl Acetate           | 4.856          | 43   | 120308     | 23.109  | ug/l # | 86       |
| 19) Methyl tert-butyl Ether  | 5.615          | 73   | 359108     | 21.931  | ug/l   | 98       |
| 20) Methylene Chloride       | 5.103          | 84   | 120617     | 20.812  | ug/l # | 81       |
| 21) trans-1,2-Dichloroethene | 5.599          | 96   | 116084     | 20.475  | ug/l   | 84       |
| 22) Diisopropyl ether        | 6.498          | 45   | 307599     | 22.091  | ug/l # | 94       |
| 23) Vinyl Acetate            | 6.434          | 43   | 1136726    | 105.862 | ug/l # | 90       |
| 24) 1,1-Dichloroethane       | 6.388          | 63   | 197496     | 21.281  | ug/l   | 96       |
| 25) 2-Butanone               | 7.335          | 43   | 313953     | 103.274 | ug/l # | 86       |
| 26) 2,2-Dichloropropane      | 7.329          | 77   | 186398     | 21.569  | ug/l   | 91       |
| 27) cis-1,2-Dichloroethene   | 7.327          | 96   | 137223     | 20.666  | ug/l   | 79       |
| 28) Bromochloromethane       | 7.659          | 49   | 75593      | 19.268  | ug/l # | 49       |
| 29) Tetrahydrofuran          | 7.683          | 42   | 195824     | 100.505 | ug/l # | 81       |
| 30) Chloroform               | 7.817          | 83   | 227502     | 20.425  | ug/l   | 99       |
| 31) Cyclohexane              | 8.099          | 56   | 186127     | 21.298  | ug/l # | 80       |
| 32) 1,1,1-Trichloroethane    | 8.016          | 97   | 214854     | 20.909  | ug/l   | 94       |
| 36) 1,1-Dichloropropene      | 8.225          | 75   | 163432     | 21.313  | ug/l   | 92       |
| 37) Ethyl Acetate            | 7.410          | 43   | 123452     | 19.325  | ug/l   | 95       |
| 38) Carbon Tetrachloride     | 8.212          | 117  | 193837     | 21.152  | ug/l   | 99       |
| 39) Methylcyclohexane        | 9.461          | 83   | 205612     | 21.514  | ug/l   | 87       |
| 40) Benzene                  | 8.461          | 78   | 493832     | 20.848  | ug/l   | 100      |

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_N\Data\VN090221\  
 Data File : VN068389.D  
 Acq On : 2 Sep 2021 16:52  
 Operator : JC/MD  
 Sample : VN0902WBS01  
 Misc : 5.00mL/MSVOA\_N/WATER  
 ALS Vial : 5 Sample Multiplier: 1

Instrument :  
 MSVOA\_N  
 ClientSampleId :  
 VN0902WBS01

Manual Integrations  
 APPROVED

MMDadoda  
 9/3/2021 5:23:55 PM

Quant Time: Sep 03 01:25:02 2021  
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_N\methods\82N083021W.M  
 Quant Title : SW846 8260  
 QLast Update : Thu Sep 02 14:59:33 2021  
 Response via : Initial Calibration

| Compound                      | R.T.   | QIon | Response | Conc    | Units  | Dev(Min) |
|-------------------------------|--------|------|----------|---------|--------|----------|
| 41) Methacrylonitrile         | 7.635  | 41   | 60328    | 20.160  | ug/l # | 85       |
| 42) 1,2-Dichloroethane        | 8.531  | 62   | 166834   | 21.015  | ug/l   | 94       |
| 43) Isopropyl Acetate         | 8.563  | 43   | 217866   | 21.081  | ug/l # | 89       |
| 44) Trichloroethene           | 9.217  | 130  | 148476   | 20.303  | ug/l   | 74       |
| 45) 1,2-Dichloropropane       | 9.488  | 63   | 115773   | 21.055  | ug/l   | 99       |
| 46) Dibromomethane            | 9.577  | 93   | 88588    | 19.776  | ug/l # | 79       |
| 47) Bromodichloromethane      | 9.762  | 83   | 175327   | 20.486  | ug/l # | 98       |
| 48) Methyl methacrylate       | 9.561  | 41   | 87263    | 19.299  | ug/l # | 81       |
| 49) 1,4-Dioxane               | 9.569  | 88   | 37084    | 347.954 | ug/l # | 84       |
| 51) 4-Methyl-2-Pentanone      | 10.330 | 43   | 633370   | 98.067  | ug/l   | 91       |
| 52) Toluene                   | 10.507 | 92   | 321614   | 20.187  | ug/l   | 99       |
| 53) t-1,3-Dichloropropene     | 10.719 | 75   | 173533   | 20.403  | ug/l   | 99       |
| 54) cis-1,3-Dichloropropene   | 10.191 | 75   | 185854   | 20.578  | ug/l # | 83       |
| 55) 1,1,2-Trichloroethane     | 10.899 | 97   | 125077   | 19.425  | ug/l   | 94       |
| 56) Ethyl methacrylate        | 10.762 | 69   | 164962   | 18.860  | ug/l # | 82       |
| 57) 1,3-Dichloropropane       | 11.046 | 76   | 191980   | 19.814  | ug/l   | 98       |
| 58) 2-Chloroethyl Vinyl ether | 10.043 | 63   | 94978    | 72.660  | ug/l # | 85       |
| 59) 2-Hexanone                | 11.087 | 43   | 447515   | 99.532  | ug/l   | 88       |
| 60) Dibromochloromethane      | 11.240 | 129  | 147785   | 19.930  | ug/l   | 99       |
| 61) 1,2-Dibromoethane         | 11.347 | 107  | 131038   | 19.265  | ug/l   | 100      |
| 64) Tetrachloroethene         | 10.979 | 164  | 146390   | 21.789  | ug/l   | 88       |
| 65) Chlorobenzene             | 11.771 | 112  | 353766   | 20.479  | ug/l   | 98       |
| 66) 1,1,1,2-Tetrachloroethane | 11.843 | 131  | 140180   | 20.797  | ug/l   | 98       |
| 67) Ethyl Benzene             | 11.848 | 91   | 592196   | 20.967  | ug/l   | 93       |
| 68) m/p-Xylenes               | 11.956 | 106  | 484336   | 42.568  | ug/l   | 79       |
| 69) o-Xylene                  | 12.283 | 106  | 229439   | 20.961  | ug/l   | 82       |
| 70) Styrene                   | 12.296 | 104  | 363297   | 21.084  | ug/l # | 91       |
| 71) Bromoform                 | 12.460 | 173  | 105645   | 17.612  | ug/l # | 99       |
| 73) Isopropylbenzene          | 12.581 | 105  | 595890   | 19.583  | ug/l   | 95       |
| 74) N-amyl acetate            | 12.387 | 43   | 148717   | 19.553  | ug/l # | 87       |
| 75) 1,1,2,2-Tetrachloroethane | 12.830 | 83   | 161557   | 17.731  | ug/l   | 98       |
| 76) 1,2,3-Trichloropropane    | 12.884 | 75   | 129690m  | 17.998  | ug/l   |          |
| 77) Bromobenzene              | 12.862 | 156  | 158616   | 18.420  | ug/l   | 56       |
| 78) n-propylbenzene           | 12.924 | 91   | 658570   | 20.431  | ug/l   | 90       |
| 79) 2-Chlorotoluene           | 13.010 | 91   | 390688   | 19.593  | ug/l   | 83       |
| 80) 1,3,5-Trimethylbenzene    | 13.061 | 105  | 503595   | 20.367  | ug/l   | 90       |
| 81) trans-1,4-Dichloro-2-b... | 12.626 | 75   | 38267    | 17.546  | ug/l   | 88       |
| 82) 4-Chlorotoluene           | 13.106 | 91   | 384788   | 19.982  | ug/l   | 86       |
| 83) tert-Butylbenzene         | 13.326 | 119  | 445175   | 19.656  | ug/l   | 83       |
| 84) 1,2,4-Trimethylbenzene    | 13.372 | 105  | 488241   | 20.186  | ug/l   | 90       |
| 85) sec-Butylbenzene          | 13.503 | 105  | 606282   | 20.353  | ug/l   | 91       |
| 86) p-Isopropyltoluene        | 13.618 | 119  | 519609   | 21.094  | ug/l   | 90       |
| 87) 1,3-Dichlorobenzene       | 13.616 | 146  | 287501   | 19.599  | ug/l   | 92       |
| 88) 1,4-Dichlorobenzene       | 13.696 | 146  | 280998   | 19.248  | ug/l   | 93       |
| 89) n-Butylbenzene            | 13.943 | 91   | 379003   | 20.411  | ug/l   | 95       |
| 90) Hexachloroethane          | 14.214 | 117  | 83846    | 17.667  | ug/l   | 75       |
| 91) 1,2-Dichlorobenzene       | 13.991 | 146  | 266396   | 19.131  | ug/l   | 92       |
| 92) 1,2-Dibromo-3-Chloropr... | 14.605 | 75   | 26624    | 18.817  | ug/l # | 36       |
| 93) 1,2,4-Trichlorobenzene    | 15.265 | 180  | 131587   | 18.823  | ug/l   | 96       |
| 94) Hexachlorobutadiene       | 15.373 | 225  | 80500    | 19.919  | ug/l   | 99       |
| 95) Naphthalene               | 15.507 | 128  | 275528   | 17.395  | ug/l   | 99       |
| 96) 1,2,3-Trichlorobenzene    | 15.702 | 180  | 124351   | 18.595  | ug/l   | 98       |

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_N\Data\VN090221\  
Data File : VN068389.D  
Acq On : 2 Sep 2021 16:52  
Operator : JC/MD  
Sample : VN0902WBS01  
Misc : 5.00mL/MSVOA\_N/WATER  
ALS Vial : 5 Sample Multiplier: 1

Instrument :  
MSVOA\_N  
ClientSampleId :  
VN0902WBS01

Manual Integrations  
APPROVED


MMDadoda  
9/3/2021 5:23:55 PM

Quant Time: Sep 03 01:25:02 2021  
Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_N\methods\82N083021W.M  
Quant Title : SW846 8260  
QLast Update : Thu Sep 02 14:59:33 2021  
Response via : Initial Calibration


| Compound | R.T. | QIon | Response | Conc | Units | Dev(Min) |
|----------|------|------|----------|------|-------|----------|
|----------|------|------|----------|------|-------|----------|

(#) = qualifier out of range (m) = manual integration (+) = signals summed







#1  
Pentafluorobenzene  
Concen: 50.000 ug/l  
RT: 8.088 min Scan# 2274  
Delta R.T. 0.000 min  
Lab File: VN068389.D  
Acq: 2 Sep 2021 16:52




Tgt Ion: 168 Resp: 649735  
Ion Ratio Lower Upper  
168 100  
99 48.6 62.9 94.3#




Instrument :  
MSVOA\_N  
ClientSampled :  
VN0902WBS01

Manual Integrations  
APPROVED



MMDadoda  
9/3/2021 5:23:55 PM




#2  
Dichlorodifluoromethane  
Concen: 19.865 ug/l  
RT: 2.075 min Scan# 32  
Delta R.T. 0.002 min  
Lab File: VN068389.D  
Acq: 2 Sep 2021 16:52



Tgt Ion: 85 Resp: 107411  
Ion Ratio Lower Upper  
85 100  
87 31.8 16.1 48.2









#3  
Chloromethane  
Concen: 17.772 ug/l  
RT: 2.309 min Scan# 119  
Delta R.T. 0.008 min  
Lab File: VN068389.D  
Acq: 2 Sep 2021 16:52

Tgt Ion: 50 Resp: 100705  
Ion Ratio Lower Upper  
50 100  
52 32.9 25.6 38.4

Instrument :  
MSVOA\_N  
ClientSampleId :  
VN0902WBS01




Manual Integrations  
APPROVED


MMDadoda  
9/3/2021 5:23:55 PM



#4  
Vinyl Chloride  
Concen: 18.655 ug/l  
RT: 2.445 min Scan# 170  
Delta R.T. 0.005 min  
Lab File: VN068389.D  
Acq: 2 Sep 2021 16:52

Tgt Ion: 62 Resp: 161181  
Ion Ratio Lower Upper  
62 100  
64 32.3 24.4 36.6






#5  
Bromomethane  
Concen: 18.165 ug/l  
RT: 2.845 min Scan# 319  
Delta R.T. 0.005 min  
Lab File: VN068389.D  
Acq: 2 Sep 2021 16:52




Instrument :  
MSVOA\_N  
ClientSampleId :  
VN0902WBS01

Manual Integrations  
APPROVED

MMDadoda  
9/3/2021 5:23:55 PM






Tgt Ion: 94 Resp: 146237  
Ion Ratio Lower Upper  
94 100  
96 95.0 73.7 110.5



#6  
Chloroethane  
Concen: 19.091 ug/l  
RT: 3.014 min Scan# 382  
Delta R.T. 0.005 min  
Lab File: VN068389.D  
Acq: 2 Sep 2021 16:52

Tgt Ion: 64 Resp: 118236  
Ion Ratio Lower Upper  
64 100  
66 32.5 24.7 37.1









#7  
Trichlorofluoromethane  
Concen: 11.978 ug/l  
RT: 3.376 min Scan# 517  
Delta R.T. 0.008 min  
Lab File: VN068389.D  
Acq: 2 Sep 2021 16:52

Instrument :  
MSVOA\_N  
ClientSampleId :  
VN0902WBS01

Tgt Ion:101 Resp: 205650  
Ion Ratio Lower Upper  
101 100  
103 66.6 51.2 76.8




Manual Integrations  
APPROVED


MMDadoda  
9/3/2021 5:23:55 PM



#8  
Diethyl Ether  
Concen: 13.160 ug/l  
RT: 3.829 min Scan# 686  
Delta R.T. 0.010 min  
Lab File: VN068389.D  
Acq: 2 Sep 2021 16:52

Tgt Ion: 74 Resp: 65081  
Ion Ratio Lower Upper  
74 100  
45 81.9 50.2 150.7









#9  
1,1,2-Trichlorotrifluoroethane  
Concen: 21.485 ug/l m  
RT: 4.224 min Scan# 833  
Delta R.T. 0.025 min  
Lab File: VN068389.D  
Acq: 2 Sep 2021 16:52

Tgt Ion:101 Resp: 116417  
Ion Ratio Lower Upper  
101 100  
85 44.5 39.6 59.4  
151 86.6 54.1 81.1#

Instrument :  
MSVOA\_N  
ClientSampleId :  
VN0902WBS01




Manual Integrations  
APPROVED


MMDadoda  
9/3/2021 5:23:55 PM



#10  
Methyl Iodide  
Concen: 23.537 ug/l  
RT: 4.427 min Scan# 909  
Delta R.T. 0.013 min  
Lab File: VN068389.D  
Acq: 2 Sep 2021 16:52

Tgt Ion:142 Resp: 162735  
Ion Ratio Lower Upper  
142 100  
127 43.7 43.1 64.7  
141 14.0 11.4 17.0








#11  
Tert butyl alcohol  
Concen: 97.881 ug/l  
RT: 5.369 min Scan# 1260  
Delta R.T. -0.024 min  
Lab File: VN068389.D  
Acq: 2 Sep 2021 16:52

Tgt Ion: 59 Resp: 84882  
Ion Ratio Lower Upper  
59 100  
57 10.7 8.7 13.1

Instrument :  
MSVOA\_N  
ClientSampleId :  
VN0902WBS01




Manual Integrations  
APPROVED

MMDadoda  
9/3/2021 5:23:55 PM



#12  
1,1-Dichloroethene  
Concen: 21.120 ug/l  
RT: 4.186 min Scan# 819  
Delta R.T. 0.013 min  
Lab File: VN068389.D  
Acq: 2 Sep 2021 16:52

Tgt Ion: 96 Resp: 106998  
Ion Ratio Lower Upper  
96 100  
61 144.1 152.6 229.0#  
98 63.2 50.9 76.3






#13  
Acrolein  
Concen: 87.679 ug/l  
RT: 4.047 min Scan# 767  
Delta R.T. 0.008 min  
Lab File: VN068389.D  
Acq: 2 Sep 2021 16:52




Instrument :  
MSVOA\_N  
ClientSampleId :  
VN0902WBS01

Manual Integrations  
APPROVED

MMDadoda  
9/3/2021 5:23:55 PM







Tgt Ion: 56 Resp: 49383  
Ion Ratio Lower Upper  
56 100  
55 74.9 57.4 86.0




#14  
Allyl chloride  
Concen: 21.847 ug/l  
RT: 4.846 min Scan# 1065  
Delta R.T. 0.014 min  
Lab File: VN068389.D  
Acq: 2 Sep 2021 16:52

Tgt Ion: 41 Resp: 139229  
Ion Ratio Lower Upper  
41 100  
39 79.0 60.9 91.3  
76 42.6 24.3 36.5#







#15  
Acrylonitrile  
Concen: 97.422 ug/l  
RT: 5.554 min Scan# 1329  
Delta R.T. -0.003 min  
Lab File: VN068389.D  
Acq: 2 Sep 2021 16:52




Tgt Ion: 53 Resp: 230989  
Ion Ratio Lower Upper  
53 100  
52 82.6 66.2 99.2  
51 37.1 30.2 45.4




Instrument :  
MSVOA\_N  
ClientSampleId :  
VN0902WBS01

Manual Integrations  
APPROVED



MMDadoda  
9/3/2021 5:23:55 PM




#16  
Acetone  
Concen: 129.177 ug/l  
RT: 4.288 min Scan# 857  
Delta R.T. -0.003 min  
Lab File: VN068389.D  
Acq: 2 Sep 2021 16:52



Tgt Ion: 43 Resp: 238937  
Ion Ratio Lower Upper  
43 100  
58 34.8 21.2 31.8#









#17  
Carbon Disulfide  
Concen: 19.693 ug/l  
RT: 4.537 min Scan# 950  
Delta R.T. 0.016 min  
Lab File: VN068389.D  
Acq: 2 Sep 2021 16:52

Tgt Ion: 76 Resp: 269020  
Ion Ratio Lower Upper  
76 100  
78 9.2 7.7 11.5

Instrument :  
MSVOA\_N  
ClientSampleId :  
VN0902WBS01




Manual Integrations  
APPROVED

MMDadoda  
9/3/2021 5:23:55 PM



#18  
Methyl Acetate  
Concen: 23.109 ug/l  
RT: 4.856 min Scan# 1069  
Delta R.T. -0.001 min  
Lab File: VN068389.D  
Acq: 2 Sep 2021 16:52

Tgt Ion: 43 Resp: 120308  
Ion Ratio Lower Upper  
43 100  
74 28.5 17.4 26.0#





#19





Methyl tert-butyl Ether  
Concen: 21.931 ug/l  
RT: 5.615 min Scan# 1352  
Delta R.T. -0.006 min  
Lab File: VN068389.D  
Acq: 2 Sep 2021 16:52

Tgt Ion: 73 Resp: 359108  
Ion Ratio Lower Upper  
73 100  
57 20.4 17.0 25.6

Instrument :  
MSVOA\_N  
ClientSampleId :  
VN0902WBS01

Manual Integrations  
APPROVED




MMDadoda  
9/3/2021 5:23:55 PM




#20

Methylene Chloride  
Concen: 20.812 ug/l  
RT: 5.103 min Scan# 1161  
Delta R.T. 0.010 min  
Lab File: VN068389.D  
Acq: 2 Sep 2021 16:52

Tgt Ion: 84 Resp: 120617  
Ion Ratio Lower Upper  
84 100  
49 104.2 110.9 166.3#  
51 32.1 32.6 49.0#  
86 65.3 51.6 77.4









#21  
trans-1,2-Dichloroethene  
Concen: 20.475 ug/l  
RT: 5.599 min Scan# 1346  
Delta R.T. 0.005 min  
Lab File: VN068389.D  
Acq: 2 Sep 2021 16:52

Tgt Ion: 96 Resp: 116084  
Ion Ratio Lower Upper  
96 100  
61 127.7 123.5 185.3  
98 63.7 48.6 72.8

Instrument :  
MSVOA\_N  
ClientSampleId :  
VN0902WBS01




Manual Integrations  
APPROVED


MMDadoda  
9/3/2021 5:23:55 PM




#22  
Diisopropyl ether  
Concen: 22.091 ug/l  
RT: 6.498 min Scan# 1681  
Delta R.T. -0.003 min  
Lab File: VN068389.D  
Acq: 2 Sep 2021 16:52

Tgt Ion: 45 Resp: 307599  
Ion Ratio Lower Upper  
45 100  
43 61.3 49.4 74.2  
87 36.3 20.9 31.3#  
59 12.0 9.0 13.4







#23  
 Vinyl Acetate  
 Concen: 105.862 ug/l  
 RT: 6.434 min Scan# 1657  
 Delta R.T. -0.002 min  
 Lab File: VN068389.D  
 Acq: 2 Sep 2021 16:52




Tgt Ion: 43 Resp: 1136726  
 Ion Ratio Lower Upper  
 43 100  
 86 12.8 7.3 10.9#




Instrument :  
 MSVOA\_N  
 ClientSampleId :  
 VN0902WBS01

Manual Integrations  
 APPROVED



MMDadoda  
 9/3/2021 5:23:55 PM




#24  
 1,1-Dichloroethane  
 Concen: 21.281 ug/l  
 RT: 6.388 min Scan# 1640  
 Delta R.T. 0.003 min  
 Lab File: VN068389.D  
 Acq: 2 Sep 2021 16:52



Tgt Ion: 63 Resp: 197496  
 Ion Ratio Lower Upper  
 63 100  
 98 7.7 3.0 9.0  
 100 4.7 1.9 5.7









#25  
2-Butanone  
Concen: 103.274 ug/l  
RT: 7.335 min Scan# 1993  
Delta R.T. -0.005 min  
Lab File: VN068389.D  
Acq: 2 Sep 2021 16:52

Tgt Ion: 43 Resp: 313953  
Ion Ratio Lower Upper  
43 100  
72 31.6 19.7 29.5#

Instrument :  
MSVOA\_N  
ClientSampleId :  
VN0902WBS01




Manual Integrations  
APPROVED


MMDadoda  
9/3/2021 5:23:55 PM



#26  
2,2-Dichloropropane  
Concen: 21.569 ug/l  
RT: 7.329 min Scan# 1991  
Delta R.T. 0.005 min  
Lab File: VN068389.D  
Acq: 2 Sep 2021 16:52

Tgt Ion: 77 Resp: 186398  
Ion Ratio Lower Upper  
77 100  
97 23.6 9.8 29.3









#27  
cis-1,2-Dichloroethene  
Concen: 20.666 ug/l  
RT: 7.327 min Scan# 1990  
Delta R.T. -0.000 min  
Lab File: VN068389.D  
Acq: 2 Sep 2021 16:52

Tgt Ion: 96 Resp: 137223  
Ion Ratio Lower Upper  
96 100  
61 130.1 0.0 338.8  
98 64.4 0.0 129.8

Instrument :  
MSVOA\_N  
ClientSampleId :  
VN0902WBS01




Manual Integrations  
APPROVED


MMDadoda  
9/3/2021 5:23:55 PM



#28  
Bromochloromethane  
Concen: 19.268 ug/l  
RT: 7.659 min Scan# 2114  
Delta R.T. 0.000 min  
Lab File: VN068389.D  
Acq: 2 Sep 2021 16:52

Tgt Ion: 49 Resp: 75593  
Ion Ratio Lower Upper  
49 100  
129 2.7 0.0 3.8  
130 102.8 49.7 74.5#









#29  
Tetrahydrofuran  
Concen: 100.505 ug/l  
RT: 7.683 min Scan# 2123  
Delta R.T. -0.008 min  
Lab File: VN068389.D  
Acq: 2 Sep 2021 16:52

Tgt Ion: 42 Resp: 195824  
Ion Ratio Lower Upper  
42 100  
72 53.6 32.8 49.2#  
71 49.9 31.2 46.8#

Instrument :  
MSVOA\_N  
ClientSampleId :  
VN0902WBS01




Manual Integrations  
APPROVED


MMDadoda  
9/3/2021 5:23:55 PM



#30  
Chloroform  
Concen: 20.425 ug/l  
RT: 7.817 min Scan# 2173  
Delta R.T. 0.000 min  
Lab File: VN068389.D  
Acq: 2 Sep 2021 16:52

Tgt Ion: 83 Resp: 227502  
Ion Ratio Lower Upper  
83 100  
85 63.3 51.5 77.3









#31  
Cyclohexane  
Concen: 21.298 ug/l  
RT: 8.099 min Scan# 2278  
Delta R.T. 0.005 min  
Lab File: VN068389.D  
Acq: 2 Sep 2021 16:52

Tgt Ion: 56 Resp: 186127  
Ion Ratio Lower Upper  
56 100  
69 39.5 27.3 40.9  
84 103.2 64.9 97.3#

Instrument :  
MSVOA\_N  
ClientSampled :  
VN0902WBS01




Manual Integrations  
APPROVED

MMDadoda  
9/3/2021 5:23:55 PM




#32  
1,1,1-Trichloroethane  
Concen: 20.909 ug/l  
RT: 8.016 min Scan# 2247  
Delta R.T. -0.000 min  
Lab File: VN068389.D  
Acq: 2 Sep 2021 16:52

Tgt Ion: 97 Resp: 214854  
Ion Ratio Lower Upper  
97 100  
99 64.2 50.2 75.2  
61 40.5 38.2 57.2







#33  
1,2-Dichloroethane-d4  
Concen: 47.733 ug/l  
RT: 8.440 min Scan# 2405  
Delta R.T. -0.000 min  
Lab File: VN068389.D  
Acq: 2 Sep 2021 16:52




Tgt Ion: 65 Resp: 316556  
Ion Ratio Lower Upper  
65 100  
67 51.8 0.0 98.6




Instrument :  
MSVOA\_N  
ClientSampleId :  
VN0902WBS01

Manual Integrations  
APPROVED



MMDadoda  
9/3/2021 5:23:55 PM




#34  
1,4-Difluorobenzene  
Concen: 50.000 ug/l  
RT: 8.968 min Scan# 2602  
Delta R.T. -0.000 min  
Lab File: VN068389.D  
Acq: 2 Sep 2021 16:52



Tgt Ion: 114 Resp: 964195  
Ion Ratio Lower Upper  
114 100  
63 17.5 0.0 49.0  
88 14.4 0.0 36.6









#35  
Dibromofluoromethane  
Concen: 47.308 ug/l  
RT: 8.024 min Scan# 2250  
Delta R.T. 0.003 min  
Lab File: VN068389.D  
Acq: 2 Sep 2021 16:52

Tgt Ion: 113 Resp: 289205  
Ion Ratio Lower Upper  
113 100  
111 101.6 83.8 125.8  
192 20.8 13.4 20.0#

Instrument :  
MSVOA\_N  
ClientSampleId :  
VN0902WBS01




Manual Integrations  
APPROVED


MMDadoda  
9/3/2021 5:23:55 PM



#36  
1,1-Dichloropropene  
Concen: 21.313 ug/l  
RT: 8.225 min Scan# 2325  
Delta R.T. 0.003 min  
Lab File: VN068389.D  
Acq: 2 Sep 2021 16:52

Tgt Ion: 75 Resp: 163432  
Ion Ratio Lower Upper  
75 100  
110 39.4 15.7 47.0  
77 30.7 24.6 37.0









#37  
Ethyl Acetate  
Concen: 19.325 ug/l  
RT: 7.410 min Scan# 2021  
Delta R.T. -0.008 min  
Lab File: VN068389.D  
Acq: 2 Sep 2021 16:52

|          |       |       |        |
|----------|-------|-------|--------|
| Tgt Ion: | 43    | Resp: | 123452 |
| Ion      | Ratio | Lower | Upper  |
| 43       | 100   |       |        |
| 61       | 15.3  | 10.6  | 16.0   |
| 70       | 12.0  | 8.1   | 12.1   |

Instrument :  
MSVOA\_N  
ClientSampleId :  
VN0902WBS01




Manual Integrations  
APPROVED


MMDadoda  
9/3/2021 5:23:55 PM



#38  
Carbon Tetrachloride  
Concen: 21.152 ug/l  
RT: 8.212 min Scan# 2320  
Delta R.T. 0.006 min  
Lab File: VN068389.D  
Acq: 2 Sep 2021 16:52

|          |       |       |        |
|----------|-------|-------|--------|
| Tgt Ion: | 117   | Resp: | 193837 |
| Ion      | Ratio | Lower | Upper  |
| 117      | 100   |       |        |
| 119      | 94.8  | 75.4  | 113.2  |
| 121      | 29.1  | 23.5  | 35.3   |





#39

Methylcyclohexane

Concen: 21.514 ug/l

RT: 9.461 min Scan# 2786

Delta R.T. 0.000 min

Lab File: VN068389.D

Acq: 2 Sep 2021 16:52

Tgt Ion: 83 Resp: 205612

| Ion | Ratio | Lower | Upper |
|-----|-------|-------|-------|
| 83  | 100   |       |       |
| 55  | 68.3  | 66.6  | 99.8  |
| 98  | 48.7  | 35.3  | 52.9  |

Instrument :





MSVOA\_N

ClientSampleId :

VN0902WBS01

Manual Integrations  
APPROVED

MMDadoda  
9/3/2021 5:23:55 PM



#40

Benzene

Concen: 20.848 ug/l

RT: 8.461 min Scan# 2413




Delta R.T. 0.000 min


Lab File: VN068389.D

Acq: 2 Sep 2021 16:52

Tgt Ion: 78 Resp: 493832

| Ion | Ratio | Lower | Upper |
|-----|-------|-------|-------|
| 78  | 100   |       |       |
| 77  | 23.4  | 18.9  | 28.3  |









#41  
Methacrylonitrile  
Concen: 20.160 ug/l  
RT: 7.635 min Scan# 2105  
Delta R.T. -0.003 min  
Lab File: VN068389.D  
Acq: 2 Sep 2021 16:52

|          |       |       |       |
|----------|-------|-------|-------|
| Tgt Ion: | 41    | Resp: | 60328 |
| Ion      | Ratio | Lower | Upper |
| 41       | 100   |       |       |
| 39       | 58.2  | 47.6  | 71.4  |
| 67       | 96.3  | 57.7  | 86.5# |
| 52       | 37.7  | 26.2  | 39.4  |

Instrument :  
MSVOA\_N  
ClientSampled :  
VN0902WBS01




Manual Integrations  
APPROVED


MMDadoda  
9/3/2021 5:23:55 PM




#42  
1,2-Dichloroethane  
Concen: 21.015 ug/l  
RT: 8.531 min Scan# 2439  
Delta R.T. -0.000 min  
Lab File: VN068389.D  
Acq: 2 Sep 2021 16:52

|          |       |       |        |
|----------|-------|-------|--------|
| Tgt Ion: | 62    | Resp: | 166834 |
| Ion      | Ratio | Lower | Upper  |
| 62       | 100   |       |        |
| 98       | 9.3   | 0.0   | 14.2   |









#43  
Isopropyl Acetate  
Concen: 21.081 ug/l  
RT: 8.563 min Scan# 2451  
Delta R.T. -0.000 min  
Lab File: VN068389.D  
Acq: 2 Sep 2021 16:52





Tgt Ion: 43 Resp: 217866  
Ion Ratio Lower Upper  
43 100  
61 28.7 19.0 28.4#  
87 16.3 9.5 14.3#



#44  
Trichloroethene  
Concen: 20.303 ug/l  
RT: 9.217 min Scan# 2695  
Delta R.T. 0.000 min  
Lab File: VN068389.D  
Acq: 2 Sep 2021 16:52




Tgt Ion:130 Resp: 148476  
Ion Ratio Lower Upper  
130 100  
95 88.3 0.0 234.4



Instrument :  
MSVOA\_N  
ClientSampled :  
VN0902WBS01

Manual Integrations  
APPROVED

MMDadoda  
9/3/2021 5:23:55 PM







#45  
1,2-Dichloropropane  
Concen: 21.055 ug/l  
RT: 9.488 min Scan# 2796  
Delta R.T. -0.003 min  
Lab File: VN068389.D  
Acq: 2 Sep 2021 16:52

Tgt Ion: 63 Resp: 115773  
Ion Ratio Lower Upper  
63 100  
65 31.0 24.3 36.5

Instrument :  
MSVOA\_N  
ClientSampleId :  
VN0902WBS01




Manual Integrations  
APPROVED


MMDadoda  
9/3/2021 5:23:55 PM



#46  
Dibromomethane  
Concen: 19.776 ug/l  
RT: 9.577 min Scan# 2829  
Delta R.T. -0.002 min  
Lab File: VN068389.D  
Acq: 2 Sep 2021 16:52


Tgt Ion: 93 Resp: 88588  
Ion Ratio Lower Upper  
93 100  
95 83.5 66.1 99.1  
174 113.2 61.0 91.4#





#47  
Bromodichloromethane  
Concen: 20.486 ug/l  
RT: 9.762 min Scan# 2898  
Delta R.T. -0.003 min  
Lab File: VN068389.D  
Acq: 2 Sep 2021 16:52




Instrument :  
MSVOA\_N  
ClientSampleId :  
VN0902WBS01




Tgt Ion: 83 Resp: 175327  
Ion Ratio Lower Upper  
83 100  
85 64.1 49.9 74.9  
127 9.2 6.1 9.1#

Manual Integrations  
APPROVED



MMDadoda  
9/3/2021 5:23:55 PM




#48  
Methyl methacrylate  
Concen: 19.299 ug/l  
RT: 9.561 min Scan# 2823  
Delta R.T. -0.000 min  
Lab File: VN068389.D  
Acq: 2 Sep 2021 16:52




Tgt Ion: 41 Resp: 87263  
Ion Ratio Lower Upper  
41 100  
69 110.2 64.6 96.8#  
39 68.5 53.3 79.9





#49  
1,4-Dioxane  
Concen: 347.954 ug/l  
RT: 9.569 min Scan# 2826  
Delta R.T. -0.002 min  
Lab File: VN068389.D  
Acq: 2 Sep 2021 16:52




Instrument :  
MSVOA\_N  
ClientSampled :  
VN0902WBS01




Tgt Ion: 88 Resp: 37084  
Ion Ratio Lower Upper  
88 100  
43 29.1 28.8 43.2  
58 59.5 60.4 90.6#

Manual Integrations  
APPROVED



MMDadoda  
9/3/2021 5:23:55 PM



#50  
Toluene-d8  
Concen: 46.622 ug/l  
RT: 10.443 min Scan# 3152  
Delta R.T. 0.000 min  
Lab File: VN068389.D  
Acq: 2 Sep 2021 16:52



Tgt Ion: 98 Resp: 1066922  
Ion Ratio Lower Upper  
98 100  
100 64.0 51.7 77.5









#51  
4-Methyl-2-Pentanone  
Concen: 98.067 ug/l  
RT: 10.330 min Scan# 3110  
Delta R.T. -0.003 min  
Lab File: VN068389.D  
Acq: 2 Sep 2021 16:52

Tgt Ion: 43 Resp: 633370  
Ion Ratio Lower Upper  
43 100  
58 40.9 28.6 43.0

Instrument :  
MSVOA\_N  
ClientSampleId :  
VN0902WBS01




Manual Integrations  
APPROVED


MMDadoda  
9/3/2021 5:23:55 PM



#52  
Toluene  
Concen: 20.187 ug/l  
RT: 10.507 min Scan# 3176  
Delta R.T. 0.000 min  
Lab File: VN068389.D  
Acq: 2 Sep 2021 16:52

Tgt Ion: 92 Resp: 321614  
Ion Ratio Lower Upper  
92 100  
91 175.4 139.6 209.4









#53  
t-1,3-Dichloropropene  
Concen: 20.403 ug/l  
RT: 10.719 min Scan# 3255  
Delta R.T. 0.000 min  
Lab File: VN068389.D  
Acq: 2 Sep 2021 16:52

Tgt Ion: 75 Resp: 173533  
Ion Ratio Lower Upper  
75 100  
77 32.2 25.1 37.7

Instrument :  
MSVOA\_N  
ClientSampleId :  
VN0902WBS01




Manual Integrations  
APPROVED


MMDadoda  
9/3/2021 5:23:55 PM




#54  
cis-1,3-Dichloropropene  
Concen: 20.578 ug/l  
RT: 10.191 min Scan# 3058  
Delta R.T. -0.000 min  
Lab File: VN068389.D  
Acq: 2 Sep 2021 16:52

Tgt Ion: 75 Resp: 185854  
Ion Ratio Lower Upper  
75 100  
77 32.4 24.6 37.0  
39 44.8 50.7 76.1#








#55  
1,1,2-Trichloroethane  
Concen: 19.425 ug/l  
RT: 10.899 min Scan# 3322  
Delta R.T. -0.003 min  
Lab File: VN068389.D  
Acq: 2 Sep 2021 16:52




Tgt Ion: 97 Resp: 125077

| Ion | Ratio | Lower | Upper |
|-----|-------|-------|-------|
| 97  | 100   |       |       |
| 83  | 81.5  | 72.0  | 108.0 |
| 85  | 53.3  | 46.8  | 70.2  |
| 99  | 62.9  | 49.8  | 74.6  |





#56  
Ethyl methacrylate  
Concen: 18.860 ug/l  
RT: 10.762 min Scan# 3271  
Delta R.T. 0.000 min  
Lab File: VN068389.D  
Acq: 2 Sep 2021 16:52



Tgt Ion: 69 Resp: 164962


| Ion | Ratio | Lower | Upper |
|-----|-------|-------|-------|
| 69  | 100   |       |       |
| 41  | 56.9  | 58.6  | 88.0# |
| 39  | 37.2  | 38.3  | 57.5# |



Instrument :  
MSVOA\_N  
ClientSampleId :  
VN0902WBS01

Manual Integrations  
APPROVED

MMDadoda  
9/3/2021 5:23:55 PM




#57  
1,3-Dichloropropane  
Concen: 19.814 ug/l  
RT: 11.046 min Scan# 3377  
Delta R.T. -0.001 min  
Lab File: VN068389.D  
Acq: 2 Sep 2021 16:52




Instrument :  
MSVOA\_N  
ClientSampleId :  
VN0902WBS01

Manual Integrations  
APPROVED

MMDadoda  
9/3/2021 5:23:55 PM







Tgt Ion: 76 Resp: 191980  
Ion Ratio Lower Upper  
76 100  
78 33.1 25.7 38.5



#58  
2-Chloroethyl Vinyl ether  
Concen: 72.660 ug/l  
RT: 10.043 min Scan# 3003  
Delta R.T. 0.000 min  
Lab File: VN068389.D  
Acq: 2 Sep 2021 16:52


Tgt Ion: 63 Resp: 94978  
Ion Ratio Lower Upper  
63 100  
106 32.3 19.8 29.8#





#59  
2-Hexanone  
Concen: 99.532 ug/l  
RT: 11.087 min Scan# 3392  
Delta R.T. -0.000 min  
Lab File: VN068389.D  
Acq: 2 Sep 2021 16:52




Instrument :  
MSVOA\_N  
ClientSampleId :  
VN0902WBS01



Tgt Ion: 43 Resp: 447515  
Ion Ratio Lower Upper  
43 100  
58 57.8 24.7 74.1




Manual Integrations  
APPROVED


MMDadoda  
9/3/2021 5:23:55 PM



#60  
Dibromochloromethane  
Concen: 19.930 ug/l  
RT: 11.240 min Scan# 3449  
Delta R.T. -0.002 min  
Lab File: VN068389.D  
Acq: 2 Sep 2021 16:52


Tgt Ion: 129 Resp: 147785  
Ion Ratio Lower Upper  
129 100  
127 76.4 38.6 115.8





#61  
1,2-Dibromoethane  
Concen: 19.265 ug/l  
RT: 11.347 min Scan# 3489  
Delta R.T. -0.000 min  
Lab File: VN068389.D  
Acq: 2 Sep 2021 16:52




Instrument :  
MSVOA\_N  
ClientSampleId :  
VN0902WBS01




Tgt Ion:107 Resp: 131038  
Ion Ratio Lower Upper  
107 100  
109 94.1 75.0 112.4

Manual Integrations  
APPROVED



MMDadoda  
9/3/2021 5:23:55 PM




#62  
4-Bromofluorobenzene  
Concen: 44.741 ug/l  
RT: 12.733 min Scan# 4006  
Delta R.T. 0.000 min  
Lab File: VN068389.D  
Acq: 2 Sep 2021 16:52




Tgt Ion: 95 Resp: 336011  
Ion Ratio Lower Upper  
95 100  
174 96.3 0.0 122.0  
176 93.0 0.0 117.6





#63  
Chlorobenzene-d5  
Concen: 50.000 ug/l  
RT: 11.746 min Scan# 3638  
Delta R.T. -0.001 min  
Lab File: VN068389.D  
Acq: 2 Sep 2021 16:52




Instrument :  
MSVOA\_N  
ClientSampleId :  
VN0902WBS01




Tgt Ion:117 Resp: 860141  
Ion Ratio Lower Upper  
117 100  
82 49.6 53.1 79.7#  
119 32.1 25.5 38.3

Manual Integrations  
APPROVED



MMDadoda  
9/3/2021 5:23:55 PM




#64  
Tetrachloroethene  
Concen: 21.789 ug/l  
RT: 10.979 min Scan# 3352  
Delta R.T. 0.000 min  
Lab File: VN068389.D  
Acq: 2 Sep 2021 16:52




Tgt Ion:164 Resp: 146390  
Ion Ratio Lower Upper  
164 100  
166 127.9 102.1 153.1  
129 88.8 86.9 130.3  
131 86.6 83.6 125.4





#65  
Chlorobenzene  
Concen: 20.479 ug/l  
RT: 11.771 min Scan# 3647  
Delta R.T. -0.002 min  
Lab File: VN068389.D  
Acq: 2 Sep 2021 16:52




Instrument :  
MSVOA\_N  
ClientSampleId :  
VN0902WBS01




Tgt Ion:112 Resp: 353766  
Ion Ratio Lower Upper  
112 100  
114 31.7 24.6 37.0

Manual Integrations  
APPROVED



MMDadoda  
9/3/2021 5:23:55 PM




#66  
1,1,1,2-Tetrachloroethane  
Concen: 20.797 ug/l  
RT: 11.843 min Scan# 3674  
Delta R.T. 0.000 min  
Lab File: VN068389.D  
Acq: 2 Sep 2021 16:52



Tgt Ion:131 Resp: 140180  
Ion Ratio Lower Upper  
131 100  
133 97.1 47.5 142.5  
119 64.3 33.1 99.3






#67  
Ethyl Benzene  
Concen: 20.967 ug/l  
RT: 11.848 min Scan# 3676  
Delta R.T. 0.000 min  
Lab File: VN068389.D  
Acq: 2 Sep 2021 16:52




Instrument :  
MSVOA\_N  
ClientSampled :  
VN0902WBS01

Manual Integrations  
APPROVED

MMDadoda  
9/3/2021 5:23:55 PM







Tgt Ion: 91 Resp: 592196  
Ion Ratio Lower Upper  
91 100  
106 32.8 23.4 35.0



#68  
m/p-Xylenes  
Concen: 42.568 ug/l  
RT: 11.956 min Scan# 3716  
Delta R.T. -0.000 min  
Lab File: VN068389.D  
Acq: 2 Sep 2021 16:52

Tgt Ion: 106 Resp: 484336  
Ion Ratio Lower Upper  
106 100  
91 191.8 180.6 270.8









#69  
o-Xylene  
Concen: 20.961 ug/l  
RT: 12.283 min Scan# 3838  
Delta R.T. -0.000 min  
Lab File: VN068389.D  
Acq: 2 Sep 2021 16:52

Instrument :  
MSVOA\_N  
ClientSampleId :  
VN0902WBS01

Tgt Ion:106 Resp: 229439  
Ion Ratio Lower Upper  
106 100  
91 206.1 118.3 354.9




Manual Integrations  
APPROVED


MMDadoda  
9/3/2021 5:23:55 PM



#70  
Styrene  
Concen: 21.084 ug/l  
RT: 12.296 min Scan# 3843  
Delta R.T. 0.000 min  
Lab File: VN068389.D  
Acq: 2 Sep 2021 16:52

Tgt Ion:104 Resp: 363297  
Ion Ratio Lower Upper  
104 100  
78 47.1 48.2 72.2#  
103 55.5 44.5 66.7









#71  
Bromoform  
Concen: 17.612 ug/l  
RT: 12.460 min Scan# 3904  
Delta R.T. -0.000 min  
Lab File: VN068389.D  
Acq: 2 Sep 2021 16:52

Tgt Ion:173 Resp: 105645  
Ion Ratio Lower Upper  
173 100  
175 48.4 23.9 71.8  
254 0.0 0.0 0.0

Instrument :  
MSVOA\_N  
ClientSampleId :  
VN0902WBS01




Manual Integrations  
APPROVED


MMDadoda  
9/3/2021 5:23:55 PM



#72  
1,4-Dichlorobenzene-d4  
Concen: 50.000 ug/l  
RT: 13.678 min Scan# 4358  
Delta R.T. 0.003 min  
Lab File: VN068389.D  
Acq: 2 Sep 2021 16:52


Tgt Ion:152 Resp: 423995  
Ion Ratio Lower Upper  
152 100  
115 53.5 36.1 108.5  
150 164.1 0.0 351.8





#73  
Isopropylbenzene  
Concen: 19.583 ug/l  
RT: 12.581 min Scan# 3949  
Delta R.T. -0.000 min  
Lab File: VN068389.D  
Acq: 2 Sep 2021 16:52




Instrument :  
MSVOA\_N  
ClientSampleId :  
VN0902WBS01




Tgt Ion:105 Resp: 595890  
Ion Ratio Lower Upper  
105 100  
120 27.7 12.5 37.5

Manual Integrations  
APPROVED



MMDadoda  
9/3/2021 5:23:55 PM




#74  
N-ethyl acetate  
Concen: 19.553 ug/l  
RT: 12.387 min Scan# 3877  
Delta R.T. -0.003 min  
Lab File: VN068389.D  
Acq: 2 Sep 2021 16:52




Tgt Ion: 43 Resp: 148717  
Ion Ratio Lower Upper  
43 100  
70 47.7 29.4 44.2#  
55 26.3 17.0 25.4#  
61 25.2 17.6 26.4





#75  
1,1,2,2-Tetrachloroethane  
Concen: 17.731 ug/l  
RT: 12.830 min Scan# 4042  
Delta R.T. 0.000 min  
Lab File: VN068389.D  
Acq: 2 Sep 2021 16:52




Instrument :  
MSVOA\_N  
ClientSampled :  
VN0902WBS01




Tgt Ion: 83 Resp: 161557  
Ion Ratio Lower Upper  
83 100  
131 11.9 4.8 14.4  
85 65.1 32.2 96.6

Manual Integrations  
APPROVED



MMDadoda  
9/3/2021 5:23:55 PM




#76  
1,2,3-Trichloropropane  
Concen: 17.998 ug/l m  
RT: 12.884 min Scan# 4062  
Delta R.T. 0.003 min  
Lab File: VN068389.D  
Acq: 2 Sep 2021 16:52




Tgt Ion: 75 Resp: 129690  
Ion Ratio Lower Upper  
75 100  
77 0.0 0.0 0.0









#77  
Bromobenzene  
Concen: 18.420 ug/l  
RT: 12.862 min Scan# 4054  
Delta R.T. 0.000 min  
Lab File: VN068389.D  
Acq: 2 Sep 2021 16:52





Tgt Ion:156 Resp: 158616  
Ion Ratio Lower Upper  
156 100  
77 162.2 136.2 408.6  
158 96.6 48.4 145.3



#78  
n-propylbenzene  
Concen: 20.431 ug/l  
RT: 12.924 min Scan# 4077  
Delta R.T. -0.000 min  
Lab File: VN068389.D  
Acq: 2 Sep 2021 16:52




Tgt Ion: 91 Resp: 658570  
Ion Ratio Lower Upper  
91 100  
120 25.7 10.4 31.4



Instrument :  
MSVOA\_N  
ClientSampleId :  
VN0902WBS01


Manual Integrations  
APPROVED

MMDadoda  
9/3/2021 5:23:55 PM



#79  
2-Chlorotoluene  
Concen: 19.593 ug/l  
RT: 13.010 min Scan# 4109  
Delta R.T. -0.000 min  
Lab File: VN068389.D  
Acq: 2 Sep 2021 16:52




Instrument :  
MSVOA\_N  
ClientSampled :  
VN0902WBS01




Tgt Ion: 91 Resp: 390688  
Ion Ratio Lower Upper  
91 100  
126 38.2 14.5 43.5

Manual Integrations  
APPROVED



MMDadoda  
9/3/2021 5:23:55 PM




#80  
1,3,5-Trimethylbenzene  
Concen: 20.367 ug/l  
RT: 13.061 min Scan# 4128  
Delta R.T. -0.000 min  
Lab File: VN068389.D  
Acq: 2 Sep 2021 16:52




Tgt Ion: 105 Resp: 503595  
Ion Ratio Lower Upper  
105 100  
120 50.8 22.2 66.6









#81  
trans-1,4-Dichloro-2-butene  
Concen: 17.546 ug/l  
RT: 12.626 min Scan# 3966  
Delta R.T. -0.003 min  
Lab File: VN068389.D  
Acq: 2 Sep 2021 16:52





Tgt Ion: 75 Resp: 38267  
Ion Ratio Lower Upper  
75 100  
53 93.2 87.6 131.4  
89 49.2 37.8 56.8



#82  
4-Chlorotoluene  
Concen: 19.982 ug/l  
RT: 13.106 min Scan# 4145  
Delta R.T. 0.000 min  
Lab File: VN068389.D  
Acq: 2 Sep 2021 16:52



Tgt Ion: 91 Resp: 384788  
Ion Ratio Lower Upper  
91 100  
126 37.4 14.8 44.4



Instrument :  
MSVOA\_N  
ClientSampleId :  
VN0902WBS01


Manual Integrations  
APPROVED

MMDadoda  
9/3/2021 5:23:55 PM



#83  
tert-Butylbenzene  
Concen: 19.656 ug/l  
RT: 13.326 min Scan# 4227  
Delta R.T. 0.000 min  
Lab File: VN068389.D  
Acq: 2 Sep 2021 16:52




Instrument :  
MSVOA\_N  
ClientSampleId :  
VN0902WBS01




Tgt Ion:119 Resp: 445175  
Ion Ratio Lower Upper  
119 100  
91 60.9 39.5 118.5  
134 27.2 12.0 36.0

Manual Integrations  
APPROVED



MMDadoda  
9/3/2021 5:23:55 PM




#84  
1,2,4-Trimethylbenzene  
Concen: 20.186 ug/l  
RT: 13.372 min Scan# 4244  
Delta R.T. 0.003 min  
Lab File: VN068389.D  
Acq: 2 Sep 2021 16:52




Tgt Ion:105 Resp: 488241  
Ion Ratio Lower Upper  
105 100  
120 47.6 20.6 61.8





#85  
 sec-Butylbenzene  
 Concen: 20.353 ug/l  
 RT: 13.503 min Scan# 4293  
 Delta R.T. 0.000 min  
 Lab File: VN068389.D  
 Acq: 2 Sep 2021 16:52




Instrument :  
 MSVOA\_N  
 ClientSampleId :  
 VN0902WBS01




Tgt Ion:105 Resp: 606282  
 Ion Ratio Lower Upper  
 105 100  
 134 21.9 9.0 27.1

Manual Integrations  
 APPROVED



MMDadoda  
 9/3/2021 5:23:55 PM




#86  
 p-Isopropyltoluene  
 Concen: 21.094 ug/l  
 RT: 13.618 min Scan# 4336  
 Delta R.T. -0.001 min  
 Lab File: VN068389.D  
 Acq: 2 Sep 2021 16:52




Tgt Ion:119 Resp: 519609  
 Ion Ratio Lower Upper  
 119 100  
 134 27.7 12.4 37.4  
 91 21.8 14.8 44.3





#87  
1,3-Dichlorobenzene  
Concen: 19.599 ug/l  
RT: 13.616 min Scan# 4335  
Delta R.T. -0.003 min  
Lab File: VN068389.D  
Acq: 2 Sep 2021 16:52




Instrument :  
MSVOA\_N  
ClientSampleId :  
VN0902WBS01




Tgt Ion:146 Resp: 287501  
Ion Ratio Lower Upper  
146 100  
111 36.3 24.3 72.8  
148 63.6 32.0 96.0

Manual Integrations  
APPROVED



MMDadoda  
9/3/2021 5:23:55 PM




#88  
1,4-Dichlorobenzene  
Concen: 19.248 ug/l  
RT: 13.696 min Scan# 4365  
Delta R.T. 0.000 min  
Lab File: VN068389.D  
Acq: 2 Sep 2021 16:52




Tgt Ion:146 Resp: 280998  
Ion Ratio Lower Upper  
146 100  
111 36.6 23.4 70.2  
148 64.5 31.8 95.3









#89  
n-Butylbenzene  
Concen: 20.411 ug/l  
RT: 13.943 min Scan# 4457  
Delta R.T. 0.000 min  
Lab File: VN068389.D  
Acq: 2 Sep 2021 16:52





Tgt Ion: 91 Resp: 379003  
Ion Ratio Lower Upper  
91 100  
92 50.9 25.7 77.1  
134 28.8 11.2 33.6



#90  
Hexachloroethane  
Concen: 17.667 ug/l  
RT: 14.214 min Scan# 4558  
Delta R.T. 0.003 min  
Lab File: VN068389.D  
Acq: 2 Sep 2021 16:52




Tgt Ion: 117 Resp: 83846  
Ion Ratio Lower Upper  
117 100  
201 87.2 33.5 100.5



Instrument :  
MSVOA\_N  
ClientSampleId :  
VN0902WBS01


Manual Integrations  
APPROVED

MMDadoda  
9/3/2021 5:23:55 PM



#91  
1,2-Dichlorobenzene  
Concen: 19.131 ug/l  
RT: 13.991 min Scan# 4475  
Delta R.T. 0.000 min  
Lab File: VN068389.D  
Acq: 2 Sep 2021 16:52




Instrument :  
MSVOA\_N  
ClientSampleId :  
VN0902WBS01




Tgt Ion:146 Resp: 266396  
Ion Ratio Lower Upper  
146 100  
111 39.1 24.9 74.6  
148 64.9 31.6 95.0

Manual Integrations  
APPROVED



MMDadoda  
9/3/2021 5:23:55 PM



#92  
1,2-Dibromo-3-Chloropropane  
Concen: 18.817 ug/l  
RT: 14.605 min Scan# 4704  
Delta R.T. 0.000 min  
Lab File: VN068389.D  
Acq: 2 Sep 2021 16:52




Tgt Ion: 75 Resp: 26624  
Ion Ratio Lower Upper  
75 100  
155 102.4 29.8 89.5#  
157 134.7 37.6 112.8#





#93  
1,2,4-Trichlorobenzene  
Concen: 18.823 ug/l  
RT: 15.265 min Scan# 4950  
Delta R.T. 0.000 min  
Lab File: VN068389.D  
Acq: 2 Sep 2021 16:52




Instrument :  
MSVOA\_N  
ClientSampleId :  
VN0902WBS01




Tgt Ion:180 Resp: 131587  
Ion Ratio Lower Upper  
180 100  
182 93.2 47.5 142.5  
145 29.2 17.2 51.6

Manual Integrations  
APPROVED



MMDadoda  
9/3/2021 5:23:55 PM




#94  
Hexachlorobutadiene  
Concen: 19.919 ug/l  
RT: 15.373 min Scan# 4990  
Delta R.T. -0.000 min  
Lab File: VN068389.D  
Acq: 2 Sep 2021 16:52




Tgt Ion:225 Resp: 80500  
Ion Ratio Lower Upper  
225 100  
223 61.6 31.4 94.0  
227 62.9 31.2 93.6





#95  
Naphthalene  
Concen: 17.395 ug/l  
RT: 15.507 min Scan# 5040  
Delta R.T. -0.000 min  
Lab File: VN068389.D  
Acq: 2 Sep 2021 16:52




Instrument :  
MSVOA\_N  
ClientSampleId :  
VN0902WBS01




Tgt Ion:128 Resp: 275528  
Ion Ratio Lower Upper  
128 100  
127 13.4 10.6 15.8  
129 11.0 8.6 12.8

Manual Integrations  
APPROVED

MMDadoda  
9/3/2021 5:23:55 PM



#96  
1,2,3-Trichlorobenzene  
Concen: 18.595 ug/l  
RT: 15.702 min Scan# 5113  
Delta R.T. 0.000 min  
Lab File: VN068389.D  
Acq: 2 Sep 2021 16:52



Tgt Ion:180 Resp: 124351  
Ion Ratio Lower Upper  
180 100  
182 95.5 47.8 143.3  
145 31.8 18.2 54.6

