

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN090222\
 Data File : VN074292.D
 Acq On : 02 Sep 2022 17:59
 Operator : JC\MD
 Sample : VN0902WBS01
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VN0902WBS01

Quant Time: Sep 03 04:40:24 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N090222W.M
 Quant Title : SW846 8260
 QLast Update : Sat Sep 03 04:35:33 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.010	168	263247	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.898	114	478161	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.680	117	442844	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.610	152	206772	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.369	65	206954	50.433	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	= 100.860%	
35) Dibromofluoromethane	7.951	113	163349	51.553	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 103.100%	
50) Toluene-d8	10.381	98	673130	53.875	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	= 107.760%	
62) 4-Bromofluorobenzene	12.669	95	217939	47.802	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 95.600%	

Target Compounds				Qvalue		
2) Dichlorodifluoromethane	2.046	85	44971	18.039	ug/l	98
3) Chloromethane	2.269	50	63835	17.749	ug/l	97
4) Vinyl Chloride	2.405	62	135172	20.197	ug/l	98
5) Bromomethane	2.781	94	97520	21.675	ug/l	88
6) Chloroethane	2.957	64	113171	21.472	ug/l	98
7) Trichlorofluoromethane	3.310	101	90192	19.067	ug/l	97
8) Diethyl Ether	3.763	74	35681	18.526	ug/l	96
9) 1,1,2-Trichlorotrifluo...	4.140	101	52935	18.454	ug/l	95
10) Methyl Iodide	4.351	142	29985	18.370	ug/l #	85
11) Tert butyl alcohol	5.281	59	67579	90.774	ug/l #	88
12) 1,1-Dichloroethene	4.116	96	48364	18.665	ug/l	92
13) Acrolein	3.981	56	47468	94.733	ug/l	98
14) Allyl chloride	4.757	41	75160	18.037	ug/l #	85
15) Acrylonitrile	5.463	53	190535	93.058	ug/l	99
16) Acetone	4.216	43	158897	92.651	ug/l	98
17) Carbon Disulfide	4.451	76	110150	17.022	ug/l	96
18) Methyl Acetate	4.775	43	91825	17.790	ug/l	99
19) Methyl tert-butyl Ether	5.522	73	200305	19.502	ug/l	94
20) Methylene Chloride	5.010	84	65106	20.251	ug/l	94
21) trans-1,2-Dichloroethene	5.516	96	55332	18.630	ug/l	90
22) Diisopropyl ether	6.404	45	192403	19.667	ug/l	95
23) Vinyl Acetate	6.345	43	822848	96.769	ug/l	99
24) 1,1-Dichloroethane	6.304	63	114746	19.472	ug/l	96
25) 2-Butanone	7.257	43	277385	95.781	ug/l	99
26) 2,2-Dichloropropane	7.245	77	93380	19.505	ug/l	99
27) cis-1,2-Dichloroethene	7.245	96	75103	19.878	ug/l	93
28) Bromochloromethane	7.575	49	40635	20.045	ug/l	86
29) Tetrahydrofuran	7.610	42	179717	97.258	ug/l	99
30) Chloroform	7.739	83	129917	19.930	ug/l	90
31) Cyclohexane	8.016	56	101230	18.470	ug/l	94
32) 1,1,1-Trichloroethane	7.934	97	109937	20.333	ug/l	98
36) 1,1-Dichloropropene	8.145	75	90005	20.014	ug/l	97
37) Ethyl Acetate	7.334	43	104905	17.836	ug/l	96
38) Carbon Tetrachloride	8.134	117	92708	20.407	ug/l	95
39) Methylcyclohexane	9.392	83	111294	20.541	ug/l	92
40) Benzene	8.387	78	282777	19.933	ug/l	99

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN090222\
 Data File : VN074292.D
 Acq On : 02 Sep 2022 17:59
 Operator : JC\MD
 Sample : VN0902WBS01
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VN0902WBS01

Quant Time: Sep 03 04:40:24 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N090222W.M
 Quant Title : SW846 8260
 QLast Update : Sat Sep 03 04:35:33 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
41) Methacrylonitrile	7.569	41	55890	21.128	ug/l	92
42) 1,2-Dichloroethane	8.463	62	101296	19.195	ug/l	99
43) Isopropyl Acetate	8.492	43	168060	19.663	ug/l	98
44) Trichloroethene	9.151	130	69300	20.526	ug/l	94
45) 1,2-Dichloropropane	9.428	63	71097	20.147	ug/l	93
46) Dibromomethane	9.516	93	52947	19.677	ug/l #	87
47) Bromodichloromethane	9.704	83	103707	21.018	ug/l	100
48) Methyl methacrylate	9.498	41	70833	18.858	ug/l #	80
49) 1,4-Dioxane	9.504	88	37286	385.701	ug/l	89
51) 4-Methyl-2-Pentanone	10.269	43	611213	102.225	ug/l	96
52) Toluene	10.439	92	193141	22.031	ug/l	98
53) t-1,3-Dichloropropene	10.657	75	107642	21.427	ug/l	91
54) cis-1,3-Dichloropropene	10.128	75	112003	20.322	ug/l #	88
55) 1,1,2-Trichloroethane	10.839	97	75657	20.114	ug/l	97
56) Ethyl methacrylate	10.698	69	121586	21.712	ug/l #	94
57) 1,3-Dichloropropane	10.980	76	132917	20.662	ug/l	99
58) 2-Chloroethyl Vinyl ether	9.975	63	288596	105.345	ug/l	95
59) 2-Hexanone	11.028	43	470351	103.941	ug/l	95
60) Dibromochloromethane	11.175	129	76718	20.552	ug/l	96
61) 1,2-Dibromoethane	11.286	107	77657	20.607	ug/l	99
64) Tetrachloroethene	10.916	164	58868	21.146	ug/l	94
65) Chlorobenzene	11.710	112	200691	20.646	ug/l	95
66) 1,1,1,2-Tetrachloroethane	11.780	131	74677	20.760	ug/l	97
67) Ethyl Benzene	11.786	91	365733	20.494	ug/l	96
68) m/p-Xylenes	11.892	106	284566	42.717	ug/l	100
69) o-Xylene	12.222	106	141388	21.265	ug/l	98
70) Styrene	12.233	104	231492	21.743	ug/l	100
71) Bromoform	12.398	173	52864	20.293	ug/l #	98
73) Isopropylbenzene	12.516	105	368323	20.342	ug/l	98
74) N-amyl acetate	12.327	43	154686	20.031	ug/l	94
75) 1,1,2,2-Tetrachloroethane	12.769	83	129702	19.306	ug/l	98
76) 1,2,3-Trichloropropane	12.816	75	157399	26.877	ug/l	44
77) Bromobenzene	12.798	156	78898	19.630	ug/l	88
78) n-propylbenzene	12.857	91	436167	21.074	ug/l	97
79) 2-Chlorotoluene	12.945	91	248377	19.547	ug/l	99
80) 1,3,5-Trimethylbenzene	12.998	105	304923	21.004	ug/l	96
81) trans-1,4-Dichloro-2-b...	12.569	75	36999	20.510	ug/l	94
82) 4-Chlorotoluene	13.039	91	236744	19.846	ug/l	98
83) tert-Butylbenzene	13.263	119	271525	20.607	ug/l	97
84) 1,2,4-Trimethylbenzene	13.304	105	302011	20.537	ug/l	97
85) sec-Butylbenzene	13.439	105	389044	21.149	ug/l	98
86) p-Isopropyltoluene	13.557	119	311498	21.737	ug/l	98
87) 1,3-Dichlorobenzene	13.551	146	156502	21.271	ug/l	98
88) 1,4-Dichlorobenzene	13.633	146	151458	20.749	ug/l	97
89) n-Butylbenzene	13.880	91	256044	20.480	ug/l	97
90) Hexachloroethane	14.151	117	53366	19.144	ug/l	85
91) 1,2-Dichlorobenzene	13.921	146	154724	20.062	ug/l	98
92) 1,2-Dibromo-3-Chloropr...	14.539	75	25341	17.376	ug/l	93
93) 1,2,4-Trichlorobenzene	15.198	180	74721	20.089	ug/l	97
94) Hexachlorobutadiene	15.298	225	35616	22.053	ug/l	97
95) Naphthalene	15.433	128	282713	19.366	ug/l	98
96) 1,2,3-Trichlorobenzene	15.627	180	77070	20.064	ug/l	99


Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN090222\
Data File : VN074292.D
Acq On : 02 Sep 2022 17:59
Operator : JC\MD
Sample : VN0902WBS01
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VN0902WBS01

Quant Time: Sep 03 04:40:24 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N090222W.M
Quant Title : SW846 8260
QLast Update : Sat Sep 03 04:35:33 2022
Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
----------	------	------	----------	------	-------	----------





(#) = qualifier out of range (m) = manual integration (+) = signals summed



#1
Pentafluorobenzene
Concen: 50.000 ug/l
RT: 8.010 min Scan# 1042
Delta R.T. -0.006 min
Lab File: VN074292.D
Acq: 02 Sep 2022 17:59




Instrument :
MSVOA_N
ClientSampleId :
VN0902WBS01


Tgt Ion:168 Resp: 263247
Ion Ratio Lower Upper
168 100
99 58.2 42.3 63.5



#2
Dichlorodifluoromethane
Concen: 18.039 ug/l
RT: 2.046 min Scan# 27
Delta R.T. -0.000 min
Lab File: VN074292.D
Acq: 02 Sep 2022 17:59


Tgt Ion: 85 Resp: 44971
Ion Ratio Lower Upper
85 100
87 30.4 15.7 47.1








#3
Chloromethane
Concen: 17.749 ug/l
RT: 2.269 min Scan# 61
Delta R.T. 0.000 min
Lab File: VN074292.D
Acq: 02 Sep 2022 17:59

Instrument :
MSVOA_N
ClientSampleId :
VN0902WBS01







Tgt Ion: 50 Resp: 63835
Ion Ratio Lower Upper
50 100
52 34.0 25.7 38.5



#4
Vinyl Chloride
Concen: 20.197 ug/l
RT: 2.405 min Scan# 88
Delta R.T. 0.001 min
Lab File: VN074292.D
Acq: 02 Sep 2022 17:59

Tgt Ion: 62 Resp: 135172
Ion Ratio Lower Upper
62 100
64 30.2 25.0 37.4









#5
Bromomethane
Concen: 21.675 ug/l
RT: 2.781 min Scan# 11
Delta R.T. 0.000 min
Lab File: VN074292.D
Acq: 02 Sep 2022 17:59




Instrument :
MSVOA_N
ClientSampleId :
VN0902WBS01


Tgt Ion: 94 Resp: 97520
Ion Ratio Lower Upper
94 100
96 83.0 76.0 114.0



#6
Chloroethane
Concen: 21.472 ug/l
RT: 2.957 min Scan# 182
Delta R.T. 0.006 min
Lab File: VN074292.D
Acq: 02 Sep 2022 17:59

Tgt Ion: 64 Resp: 113171
Ion Ratio Lower Upper
64 100
66 31.6 26.2 39.4









#7
 Trichlorofluoromethane
 Concen: 19.067 ug/l
 RT: 3.310 min Scan# 243
 Delta R.T. 0.000 min
 Lab File: VN074292.D
 Acq: 02 Sep 2022 17:59




Instrument :
 MSVOA_N
 ClientSampleId :
 VN0902WBS01


Tgt Ion:101 Resp: 90192
 Ion Ratio Lower Upper
 101 100
 103 62.1 51.5 77.3



#8
 Diethyl Ether
 Concen: 18.526 ug/l
 RT: 3.763 min Scan# 319
 Delta R.T. 0.006 min
 Lab File: VN074292.D
 Acq: 02 Sep 2022 17:59

Tgt Ion: 74 Resp: 35681
 Ion Ratio Lower Upper
 74 100
 45 82.1 42.9 128.5









#9
1,1,2-Trichlorotrifluoroethane
Concen: 18.454 ug/l
RT: 4.140 min Scan# 383
Delta R.T. 0.006 min
Lab File: VN074292.D
Acq: 02 Sep 2022 17:59




Instrument :
MSVOA_N
ClientSampleId :
VN0902WBS01


Tgt Ion:101 Resp: 52935
Ion Ratio Lower Upper
101 100
85 46.6 34.7 52.1
151 73.6 63.0 94.4



#10
Methyl Iodide
Concen: 18.370 ug/l
RT: 4.351 min Scan# 419
Delta R.T. 0.005 min
Lab File: VN074292.D
Acq: 02 Sep 2022 17:59

Tgt Ion:142 Resp: 29985
Ion Ratio Lower Upper
142 100
127 48.4 31.3 46.9#
141 8.2 11.2 16.8#





#11

Tert butyl alcohol

Concen: 90.774 ug/l

RT: 5.281 min Scan# 51

Delta R.T. -0.006 min

Lab File: VN074292.D

Acq: 02 Sep 2022 17:59

Instrument :

MSVOA_N

ClientSampleId :





VN0902WBS01

Tgt Ion: 59 Resp: 67579

Ion Ratio Lower Upper

59 100

57 5.9 8.3 12.5#



#12

1,1-Dichloroethene

Concen: 18.665 ug/l

RT: 4.116 min Scan# 379

Delta R.T. 0.006 min

Lab File: VN074292.D

Acq: 02 Sep 2022 17:59




Tgt Ion: 96 Resp: 48364


Ion Ratio Lower Upper

96 100

61 149.1 127.6 191.4

98 69.9 51.2 76.8









#13
Acrolein
Concen: 94.733 ug/l
RT: 3.981 min Scan# 356
Delta R.T. 0.006 min
Lab File: VN074292.D
Acq: 02 Sep 2022 17:59




Instrument :
MSVOA_N
ClientSampleId :
VN0902WBS01


Tgt Ion: 56 Resp: 47468
Ion Ratio Lower Upper
56 100
55 66.8 54.8 82.2



#14
Allyl chloride
Concen: 18.037 ug/l
RT: 4.757 min Scan# 488
Delta R.T. 0.000 min
Lab File: VN074292.D
Acq: 02 Sep 2022 17:59

Tgt Ion: 41 Resp: 75160
Ion Ratio Lower Upper
41 100
39 82.2 53.2 79.8#
76 41.3 29.9 44.9









#15
Acrylonitrile
Concen: 93.058 ug/l
RT: 5.463 min Scan# 610
Delta R.T. -0.006 min
Lab File: VN074292.D
Acq: 02 Sep 2022 17:59




Instrument :
MSVOA_N
ClientSampleId :
VN0902WBS01


Tgt Ion: 53 Resp: 190535
Ion Ratio Lower Upper
53 100
52 82.2 66.4 99.6
51 36.6 28.5 42.7



#16
Acetone
Concen: 92.651 ug/l
RT: 4.216 min Scan# 396
Delta R.T. -0.006 min
Lab File: VN074292.D
Acq: 02 Sep 2022 17:59

Tgt Ion: 43 Resp: 158897
Ion Ratio Lower Upper
43 100
58 33.8 26.0 39.0









#17
Carbon Disulfide
Concen: 17.022 ug/l
RT: 4.451 min Scan# 41
Delta R.T. 0.000 min
Lab File: VN074292.D
Acq: 02 Sep 2022 17:59




Instrument :
MSVOA_N
ClientSampleId :
VN0902WBS01


Tgt Ion: 76 Resp: 110150
Ion Ratio Lower Upper
76 100
78 10.5 7.2 10.8



#18
Methyl Acetate
Concen: 17.790 ug/l
RT: 4.775 min Scan# 491
Delta R.T. -0.000 min
Lab File: VN074292.D
Acq: 02 Sep 2022 17:59

Tgt Ion: 43 Resp: 91825
Ion Ratio Lower Upper
43 100
74 27.4 21.6 32.4









#19
Methyl tert-butyl Ether
Concen: 19.502 ug/l
RT: 5.522 min Scan# 618
Delta R.T. -0.012 min
Lab File: VN074292.D
Acq: 02 Sep 2022 17:59




Instrument :
MSVOA_N
ClientSampleId :
VN0902WBS01


Tgt Ion: 73 Resp: 200305
Ion Ratio Lower Upper
73 100
57 19.0 17.6 26.4



#20
Methylene Chloride
Concen: 20.251 ug/l
RT: 5.010 min Scan# 531
Delta R.T. 0.000 min
Lab File: VN074292.D
Acq: 02 Sep 2022 17:59

Tgt Ion: 84 Resp: 65106
Ion Ratio Lower Upper
84 100
49 120.3 88.3 132.5
51 31.9 28.2 42.2
86 63.4 50.6 76.0









#21
trans-1,2-Dichloroethene
Concen: 18.630 ug/l
RT: 5.516 min Scan# 617
Delta R.T. 0.000 min
Lab File: VN074292.D
Acq: 02 Sep 2022 17:59




Instrument :
MSVOA_N
ClientSampleId :
VN0902WBS01


Tgt Ion: 96 Resp: 55332
Ion Ratio Lower Upper
96 100
61 116.1 105.8 158.8
98 62.4 50.3 75.5



#22
Diisopropyl ether
Concen: 19.667 ug/l
RT: 6.404 min Scan# 768
Delta R.T. -0.012 min
Lab File: VN074292.D
Acq: 02 Sep 2022 17:59

Tgt Ion: 45 Resp: 192403
Ion Ratio Lower Upper
45 100
43 63.9 46.6 70.0
87 31.6 24.3 36.5
59 12.8 9.6 14.4









#23
 Vinyl Acetate
 Concen: 96.769 ug/l
 RT: 6.345 min Scan# 758
 Delta R.T. -0.006 min
 Lab File: VN074292.D
 Acq: 02 Sep 2022 17:59




Instrument :
 MSVOA_N
 ClientSampleId :
 VN0902WBS01


Tgt Ion: 43 Resp: 822848
 Ion Ratio Lower Upper
 43 100
 86 12.0 10.0 15.0



#24
 1,1-Dichloroethane
 Concen: 19.472 ug/l
 RT: 6.304 min Scan# 751
 Delta R.T. 0.000 min
 Lab File: VN074292.D
 Acq: 02 Sep 2022 17:59

Tgt Ion: 63 Resp: 114746
 Ion Ratio Lower Upper
 63 100
 98 8.6 3.5 10.4
 100 3.8 2.3 6.8





#25

2-Butanone

Concen: 95.781 ug/l

RT: 7.257 min Scan# 913

Delta R.T. 0.000 min

Lab File: VN074292.D

Acq: 02 Sep 2022 17:59

Instrument :

MSVOA_N

ClientSampleId :





VN0902WBS01

Tgt Ion: 43 Resp: 277385

Ion Ratio Lower Upper

43 100

72 30.2 23.7 35.5



#26

2,2-Dichloropropane

Concen: 19.505 ug/l

RT: 7.245 min Scan# 911

Delta R.T. 0.000 min

Lab File: VN074292.D




Acq: 02 Sep 2022 17:59


Tgt Ion: 77 Resp: 93380

Ion Ratio Lower Upper

77 100

97 22.2 11.4 34.2





#27

cis-1,2-Dichloroethene

Concen: 19.878 ug/l

RT: 7.245 min Scan# 911

Delta R.T. 0.000 min

Lab File: VN074292.D

Acq: 02 Sep 2022 17:59

Instrument :

MSVOA_N

ClientSampleId :

VN0902WBS01





Tgt Ion: 96 Resp: 75103

Ion Ratio Lower Upper

96 100

61 132.1 0.0 280.6

98 58.7 0.0 128.4



#28

Bromochloromethane

Concen: 20.045 ug/l

RT: 7.575 min Scan# 967

Delta R.T. -0.011 min

Lab File: VN074292.D

Acq: 02 Sep 2022 17:59




Tgt Ion: 49 Resp: 40635


Ion Ratio Lower Upper

49 100

129 1.3 0.0 4.6

130 74.0 69.7 104.5





#29

Tetrahydrofuran

Concen: 97.258 ug/l

RT: 7.610 min Scan# 91

Delta R.T. 0.000 min

Lab File: VN074292.D

Acq: 02 Sep 2022 17:59

Instrument :

MSVOA_N

ClientSampleId :

VN0902WBS01





Tgt Ion: 42 Resp: 179717

Ion Ratio Lower Upper

42 100

72 50.4 39.7 59.5

71 46.6 36.6 55.0



#30

Chloroform

Concen: 19.930 ug/l

RT: 7.739 min Scan# 995

Delta R.T. -0.006 min

Lab File: VN074292.D




Acq: 02 Sep 2022 17:59


Tgt Ion: 83 Resp: 129917

Ion Ratio Lower Upper

83 100

85 57.3 51.8 77.8









#31
Cyclohexane
Concen: 18.470 ug/l
RT: 8.016 min Scan# 1042
Delta R.T. -0.006 min
Lab File: VN074292.D
Acq: 02 Sep 2022 17:59




Instrument :
MSVOA_N
ClientSampleId :
VN0902WBS01


Tgt Ion: 56 Resp: 101230
Ion Ratio Lower Upper
56 100
69 37.9 27.8 41.8
84 101.6 76.9 115.3



#32
1,1,1-Trichloroethane
Concen: 20.333 ug/l
RT: 7.934 min Scan# 1028
Delta R.T. -0.011 min
Lab File: VN074292.D
Acq: 02 Sep 2022 17:59

Tgt Ion: 97 Resp: 109937
Ion Ratio Lower Upper
97 100
99 62.1 50.6 76.0
61 47.2 36.0 54.0





#33

1,2-Dichloroethane-d4

Concen: 50.433 ug/l

RT: 8.369 min Scan# 1102

Delta R.T. -0.000 min

Lab File: VN074292.D

Acq: 02 Sep 2022 17:59

Instrument :

MSVOA_N

ClientSampleId :





VN0902WBS01

Tgt Ion: 65 Resp: 206954

Ion Ratio Lower Upper

65 100

67 52.9 0.0 104.8



#34

1,4-Difluorobenzene

Concen: 50.000 ug/l

RT: 8.898 min Scan# 1192

Delta R.T. -0.006 min

Lab File: VN074292.D

Acq: 02 Sep 2022 17:59




Tgt Ion: 114 Resp: 478161


Ion Ratio Lower Upper

114 100

63 20.1 0.0 40.0

88 15.2 0.0 32.6





#35

Dibromofluoromethane

Concen: 51.553 ug/l

RT: 7.951 min Scan# 1031

Delta R.T. 0.000 min

Lab File: VN074292.D

Acq: 02 Sep 2022 17:59

Instrument :

MSVOA_N

ClientSampleId :

VN0902WBS01





Tgt Ion:113 Resp: 163349

Ion Ratio Lower Upper

113 100

111 101.7 81.4 122.0

192 18.5 17.0 25.6



#36

1,1-Dichloropropene

Concen: 20.014 ug/l

RT: 8.145 min Scan# 1064

Delta R.T. -0.012 min

Lab File: VN074292.D

Acq: 02 Sep 2022 17:59




Tgt Ion: 75 Resp: 90005


Ion Ratio Lower Upper

75 100

110 34.0 18.3 54.8

77 30.4 25.1 37.7









#37
Ethyl Acetate
Concen: 17.836 ug/l
RT: 7.334 min Scan# 91
Delta R.T. -0.000 min
Lab File: VN074292.D
Acq: 02 Sep 2022 17:59




Instrument :
MSVOA_N
ClientSampleId :
VN0902WBS01


Tgt Ion: 43 Resp: 104905
Ion Ratio Lower Upper
43 100
61 16.2 11.8 17.6
70 9.4 9.3 13.9



#38
Carbon Tetrachloride
Concen: 20.407 ug/l
RT: 8.134 min Scan# 1062
Delta R.T. 0.001 min
Lab File: VN074292.D
Acq: 02 Sep 2022 17:59

Tgt Ion: 117 Resp: 92708
Ion Ratio Lower Upper
117 100
119 100.7 75.8 113.8
121 29.2 24.6 37.0





#39

Methylcyclohexane

Concen: 20.541 ug/l

RT: 9.392 min Scan# 111294

Delta R.T. 0.000 min

Lab File: VN074292.D

Acq: 02 Sep 2022 17:59

Instrument :

MSVOA_N

ClientSampleId :

VN0902WBS01





Tgt Ion: 83 Resp: 111294

Ion Ratio Lower Upper

83 100

55 62.4 57.8 86.8

98 51.7 39.4 59.2



#40

Benzene

Concen: 19.933 ug/l

RT: 8.387 min Scan# 1105

Delta R.T. -0.005 min

Lab File: VN074292.D




Acq: 02 Sep 2022 17:59


Tgt Ion: 78 Resp: 282777

Ion Ratio Lower Upper

78 100


77 23.7 19.4 29.0








#41
Methacrylonitrile
Concen: 21.128 ug/l
RT: 7.569 min Scan# 91
Delta R.T. -0.000 min
Lab File: VN074292.D
Acq: 02 Sep 2022 17:59

Instrument :
MSVOA_N
ClientSampleId :
VN0902WBS01







Tgt Ion: 41 Resp: 55890
Ion Ratio Lower Upper
41 100
39 55.9 48.6 73.0
67 80.2 73.4 110.2
52 35.7 27.7 41.5



#42
1,2-Dichloroethane
Concen: 19.195 ug/l
RT: 8.463 min Scan# 1118
Delta R.T. -0.000 min
Lab File: VN074292.D
Acq: 02 Sep 2022 17:59

Tgt Ion: 62 Resp: 101296
Ion Ratio Lower Upper
62 100
98 10.1 0.0 19.4





#43

Isopropyl Acetate

Concen: 19.663 ug/l

RT: 8.492 min Scan# 1123

Delta R.T. 0.000 min

Lab File: VN074292.D

Acq: 02 Sep 2022 17:59

Instrument :

MSVOA_N

ClientSampleId :

VN0902WBS01





Tgt Ion: 43 Resp: 168060

Ion Ratio Lower Upper

43 100

61 26.3 21.9 32.9

87 14.3 12.0 18.0



#44

Trichloroethene

Concen: 20.526 ug/l

RT: 9.151 min Scan# 1235

Delta R.T. 0.000 min

Lab File: VN074292.D




Acq: 02 Sep 2022 17:59


Tgt Ion:130 Resp: 69300

Ion Ratio Lower Upper

130 100

95 87.4 0.0 186.8





#45

1,2-Dichloropropane

Concen: 20.147 ug/l

RT: 9.428 min Scan# 11

Delta R.T. -0.000 min

Lab File: VN074292.D

Acq: 02 Sep 2022 17:59

Instrument :

MSVOA_N

ClientSampleId :





VN0902WBS01

Tgt Ion: 63 Resp: 71097

Ion Ratio Lower Upper

63 100

65 34.8 24.9 37.3



#46

Dibromomethane

Concen: 19.677 ug/l

RT: 9.516 min Scan# 1297

Delta R.T. -0.000 min

Lab File: VN074292.D

Acq: 02 Sep 2022 17:59




Tgt Ion: 93 Resp: 52947


Ion Ratio Lower Upper

93 100

95 85.0 66.9 100.3

174 84.6 86.6 130.0#





#47

Bromodichloromethane

Concen: 21.018 ug/l

RT: 9.704 min Scan# 1

Delta R.T. 0.000 min

Lab File: VN074292.D

Acq: 02 Sep 2022 17:59

Instrument :

MSVOA_N

ClientSampleId :

VN0902WBS01





Tgt Ion: 83 Resp: 103707

Ion Ratio Lower Upper

83 100

85 62.8 50.5 75.7

127 8.8 6.8 10.2



#48

Methyl methacrylate

Concen: 18.858 ug/l

RT: 9.498 min Scan# 1294

Delta R.T. 0.000 min

Lab File: VN074292.D

Acq: 02 Sep 2022 17:59




Tgt Ion: 41 Resp: 70833


Ion Ratio Lower Upper

41 100

69 103.8 74.7 112.1

39 82.2 45.6 68.4#





#49

1,4-Dioxane

Concen: 385.701 ug/l

RT: 9.504 min Scan# 1295

Delta R.T. 0.000 min

Lab File: VN074292.D

Acq: 02 Sep 2022 17:59

Instrument :

MSVOA_N

ClientSampleId :

VN0902WBS01





Tgt Ion: 88 Resp: 37286

Ion Ratio Lower Upper

88 100

43 25.1 20.2 30.4

58 55.8 54.2 81.4



#50

Toluene-d8

Concen: 53.875 ug/l

RT: 10.381 min Scan# 1444

Delta R.T. 0.000 min

Lab File: VN074292.D




Acq: 02 Sep 2022 17:59


Tgt Ion: 98 Resp: 673130

Ion Ratio Lower Upper

98 100

100 63.9 52.6 78.8









#51
4-Methyl-2-Pentanone
Concen: 102.225 ug/l
RT: 10.269 min Scan# 1425
Delta R.T. -0.006 min
Lab File: VN074292.D
Acq: 02 Sep 2022 17:59




Instrument :
MSVOA_N
ClientSampleId :
VN0902WBS01


Tgt Ion: 43 Resp: 611213
Ion Ratio Lower Upper
43 100
58 39.8 34.1 51.1



#52
Toluene
Concen: 22.031 ug/l
RT: 10.439 min Scan# 1454
Delta R.T. -0.006 min
Lab File: VN074292.D
Acq: 02 Sep 2022 17:59

Tgt Ion: 92 Resp: 193141
Ion Ratio Lower Upper
92 100
91 170.6 138.6 207.8









#53
 t-1,3-Dichloropropene
 Concen: 21.427 ug/l
 RT: 10.657 min Scan# 1491
 Delta R.T. -0.006 min
 Lab File: VN074292.D
 Acq: 02 Sep 2022 17:59




Instrument :
 MSVOA_N
 ClientSampleId :
 VN0902WBS01


Tgt Ion: 75 Resp: 107642
 Ion Ratio Lower Upper
 75 100
 77 27.2 25.6 38.4



#54
 cis-1,3-Dichloropropene
 Concen: 20.322 ug/l
 RT: 10.128 min Scan# 1401
 Delta R.T. -0.000 min
 Lab File: VN074292.D
 Acq: 02 Sep 2022 17:59

Tgt Ion: 75 Resp: 112003
 Ion Ratio Lower Upper
 75 100
 77 36.7 25.2 37.8
 39 53.8 35.7 53.5#









#55
1,1,2-Trichloroethane
Concen: 20.114 ug/l
RT: 10.839 min Scan# 1522
Delta R.T. 0.000 min
Lab File: VN074292.D
Acq: 02 Sep 2022 17:59




Instrument :
MSVOA_N
ClientSampleId :
VN0902WBS01


Tgt Ion:	97	Resp:	75657
Ion	Ratio	Lower	Upper
97	100		
83	78.1	66.2	99.4
85	52.4	43.9	65.9
99	63.2	50.6	76.0



#56
Ethyl methacrylate
Concen: 21.712 ug/l
RT: 10.698 min Scan# 1498
Delta R.T. -0.006 min
Lab File: VN074292.D
Acq: 02 Sep 2022 17:59

Tgt Ion:	69	Resp:	121586
Ion	Ratio	Lower	Upper
69	100		
41	61.0	47.8	71.6
39	42.6	27.3	40.9





#57

1,3-Dichloropropane

Concen: 20.662 ug/l

RT: 10.980 min Scan# 1546

Delta R.T. -0.006 min

Lab File: VN074292.D

Acq: 02 Sep 2022 17:59

Instrument :

MSVOA_N

ClientSampleId :





VN0902WBS01

Tgt Ion: 76 Resp: 132917

Ion Ratio Lower Upper

76 100

78 32.2 26.2 39.2



#58

2-Chloroethyl Vinyl ether

Concen: 105.345 ug/l

RT: 9.975 min Scan# 1375

Delta R.T. -0.005 min

Lab File: VN074292.D




Acq: 02 Sep 2022 17:59


Tgt Ion: 63 Resp: 288596

Ion Ratio Lower Upper

63 100

106 28.1 24.6 36.8









#59
2-Hexanone
Concen: 103.941 ug/l
RT: 11.028 min Scan# 11
Delta R.T. 0.001 min
Lab File: VN074292.D
Acq: 02 Sep 2022 17:59




Instrument :
MSVOA_N
ClientSampleId :
VN0902WBS01


Tgt Ion: 43 Resp: 470351
Ion Ratio Lower Upper
43 100
58 55.9 29.7 89.1



#60
Dibromochloromethane
Concen: 20.552 ug/l
RT: 11.175 min Scan# 1579
Delta R.T. -0.005 min
Lab File: VN074292.D
Acq: 02 Sep 2022 17:59

Tgt Ion: 129 Resp: 76718
Ion Ratio Lower Upper
129 100
127 80.9 38.8 116.4









#61
1,2-Dibromoethane
Concen: 20.607 ug/l
RT: 11.286 min Scan# 1598
Delta R.T. 0.000 min
Lab File: VN074292.D
Acq: 02 Sep 2022 17:59




Instrument :
MSVOA_N
ClientSampleId :
VN0902WBS01


Tgt Ion:107 Resp: 77657
Ion Ratio Lower Upper
107 100
109 94.8 75.0 112.6



#62
4-Bromofluorobenzene
Concen: 47.802 ug/l
RT: 12.669 min Scan# 1833
Delta R.T. -0.000 min
Lab File: VN074292.D
Acq: 02 Sep 2022 17:59

Tgt Ion: 95 Resp: 217939
Ion Ratio Lower Upper
95 100
174 78.4 0.0 167.6
176 73.8 0.0 163.4









#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 11.680 min Scan# 117
Delta R.T. -0.006 min
Lab File: VN074292.D
Acq: 02 Sep 2022 17:59




Instrument :
MSVOA_N
ClientSampleId :
VN0902WBS01


Tgt Ion:117 Resp: 442844
Ion Ratio Lower Upper
117 100
82 59.2 44.2 66.2
119 33.3 25.4 38.0



#64
Tetrachloroethene
Concen: 21.146 ug/l
RT: 10.916 min Scan# 1535
Delta R.T. -0.000 min
Lab File: VN074292.D
Acq: 02 Sep 2022 17:59

Tgt Ion:164 Resp: 58868
Ion Ratio Lower Upper
164 100
166 122.0 104.2 156.2
129 84.9 73.9 110.9
131 84.7 69.4 104.2









#65
Chlorobenzene
Concen: 20.646 ug/l
RT: 11.710 min Scan# 1670
Delta R.T. -0.000 min
Lab File: VN074292.D
Acq: 02 Sep 2022 17:59




Instrument :
MSVOA_N
ClientSampleId :
VN0902WBS01


Tgt Ion:112 Resp: 200691
Ion Ratio Lower Upper
112 100
114 34.7 25.4 38.2



#66
1,1,1,2-Tetrachloroethane
Concen: 20.760 ug/l
RT: 11.780 min Scan# 1682
Delta R.T. 0.000 min
Lab File: VN074292.D
Acq: 02 Sep 2022 17:59

Tgt Ion:131 Resp: 74677
Ion Ratio Lower Upper
131 100
133 92.3 47.8 143.4
119 68.4 33.4 100.2





#67

Ethyl Benzene

Concen: 20.494 ug/l

RT: 11.786 min Scan# 1683

Delta R.T. 0.000 min

Lab File: VN074292.D

Acq: 02 Sep 2022 17:59

Instrument :

MSVOA_N

ClientSampleId :





VN0902WBS01

Tgt Ion: 91 Resp: 365733

Ion Ratio Lower Upper

91 100

106 29.0 24.9 37.3



#68

m/p-Xylenes

Concen: 42.717 ug/l

RT: 11.892 min Scan# 1701

Delta R.T. 0.000 min

Lab File: VN074292.D




Acq: 02 Sep 2022 17:59


Tgt Ion: 106 Resp: 284566

Ion Ratio Lower Upper

106 100

91 199.1 158.9 238.3









#69
o-Xylene
Concen: 21.265 ug/l
RT: 12.222 min Scan# 1756
Delta R.T. 0.001 min
Lab File: VN074292.D
Acq: 02 Sep 2022 17:59




Instrument :
MSVOA_N
ClientSampleId :
VN0902WBS01


Tgt Ion:106 Resp: 141388
Ion Ratio Lower Upper
106 100
91 214.4 106.0 318.0



#70
Styrene
Concen: 21.743 ug/l
RT: 12.233 min Scan# 1759
Delta R.T. 0.000 min
Lab File: VN074292.D
Acq: 02 Sep 2022 17:59

Tgt Ion:104 Resp: 231492
Ion Ratio Lower Upper
104 100
78 49.4 39.5 59.3
103 55.4 44.0 66.0









#71
Bromoform
Concen: 20.293 ug/l
RT: 12.398 min Scan# 1787
Delta R.T. 0.000 min
Lab File: VN074292.D
Acq: 02 Sep 2022 17:59




Instrument :
MSVOA_N
ClientSampleId :
VN0902WBS01


Tgt Ion:173 Resp: 52864
Ion Ratio Lower Upper
173 100
175 50.4 24.6 73.6
254 0.0 0.1 0.1#



#72
1,4-Dichlorobenzene-d4
Concen: 50.000 ug/l
RT: 13.610 min Scan# 1993
Delta R.T. -0.000 min
Lab File: VN074292.D
Acq: 02 Sep 2022 17:59

Tgt Ion:152 Resp: 206772
Ion Ratio Lower Upper
152 100
115 61.0 28.9 86.7
150 164.5 0.0 356.2









#73
Isopropylbenzene
Concen: 20.342 ug/l
RT: 12.516 min Scan# 11
Delta R.T. -0.000 min
Lab File: VN074292.D
Acq: 02 Sep 2022 17:59




Instrument :
MSVOA_N
ClientSampleId :
VN0902WBS01


Tgt Ion:105 Resp: 368323
Ion Ratio Lower Upper
105 100
120 26.0 13.5 40.4



#74
N-ethyl acetate
Concen: 20.031 ug/l
RT: 12.327 min Scan# 1775
Delta R.T. 0.000 min
Lab File: VN074292.D
Acq: 02 Sep 2022 17:59

Tgt Ion: 43 Resp: 154686
Ion Ratio Lower Upper
43 100
70 44.2 39.0 58.4
55 23.5 21.1 31.7
61 24.6 21.1 31.7









#75
 1,1,2,2-Tetrachloroethane
 Concen: 19.306 ug/l
 RT: 12.769 min Scan# 1850
 Delta R.T. -0.000 min
 Lab File: VN074292.D
 Acq: 02 Sep 2022 17:59




Instrument :
 MSVOA_N
 ClientSampleId :
 VN0902WBS01


Tgt Ion: 83 Resp: 129702
 Ion Ratio Lower Upper
 83 100
 131 9.0 5.3 15.9
 85 65.8 32.3 96.9



#76
 1,2,3-Trichloropropane
 Concen: 26.877 ug/l
 RT: 12.816 min Scan# 1858
 Delta R.T. -0.005 min
 Lab File: VN074292.D
 Acq: 02 Sep 2022 17:59

Tgt Ion: 75 Resp: 157399
 Ion Ratio Lower Upper
 75 100
 77 113.3 98.5 295.5









#77
Bromobenzene
Concen: 19.630 ug/l
RT: 12.798 min Scan# 1855
Delta R.T. 0.000 min
Lab File: VN074292.D
Acq: 02 Sep 2022 17:59




Instrument :
MSVOA_N
ClientSampleId :
VN0902WBS01


Tgt Ion:156 Resp: 78898
Ion Ratio Lower Upper
156 100
77 226.0 100.5 301.5
158 101.0 48.9 146.6



#78
n-propylbenzene
Concen: 21.074 ug/l
RT: 12.857 min Scan# 1865
Delta R.T. -0.006 min
Lab File: VN074292.D
Acq: 02 Sep 2022 17:59

Tgt Ion: 91 Resp: 436167
Ion Ratio Lower Upper
91 100
120 22.3 11.9 35.5









#79
2-Chlorotoluene
Concen: 19.547 ug/l
RT: 12.945 min Scan# 1880
Delta R.T. 0.000 min
Lab File: VN074292.D
Acq: 02 Sep 2022 17:59




Instrument :
MSVOA_N
ClientSampleId :
VN0902WBS01


Tgt Ion: 91 Resp: 248377
Ion Ratio Lower Upper
91 100
126 34.3 17.3 51.9



#80
1,3,5-Trimethylbenzene
Concen: 21.004 ug/l
RT: 12.998 min Scan# 1889
Delta R.T. -0.000 min
Lab File: VN074292.D
Acq: 02 Sep 2022 17:59

Tgt Ion: 105 Resp: 304923
Ion Ratio Lower Upper
105 100
120 47.3 24.9 74.7









#81
 trans-1,4-Dichloro-2-butene
 Concen: 20.510 ug/l
 RT: 12.569 min Scan# 1816
 Delta R.T. 0.006 min
 Lab File: VN074292.D
 Acq: 02 Sep 2022 17:59




Instrument :
 MSVOA_N
 ClientSampleId :
 VN0902WBS01


Tgt Ion: 75 Resp: 36999
 Ion Ratio Lower Upper
 75 100
 53 104.7 78.0 117.0
 89 43.7 37.2 55.8



#82
 4-Chlorotoluene
 Concen: 19.846 ug/l
 RT: 13.039 min Scan# 1896
 Delta R.T. -0.006 min
 Lab File: VN074292.D
 Acq: 02 Sep 2022 17:59

Tgt Ion: 91 Resp: 236744
 Ion Ratio Lower Upper
 91 100
 126 35.6 17.2 51.6









#83
 tert-Butylbenzene
 Concen: 20.607 ug/l
 RT: 13.263 min Scan# 1934
 Delta R.T. -0.000 min
 Lab File: VN074292.D
 Acq: 02 Sep 2022 17:59




Instrument :
 MSVOA_N
 ClientSampleId :
 VN0902WBS01


Tgt Ion:119 Resp: 271525
 Ion Ratio Lower Upper
 119 100
 91 65.4 31.6 94.8
 134 26.0 13.7 41.1



#84
 1,2,4-Trimethylbenzene
 Concen: 20.537 ug/l
 RT: 13.304 min Scan# 1941
 Delta R.T. -0.006 min
 Lab File: VN074292.D
 Acq: 02 Sep 2022 17:59

Tgt Ion:105 Resp: 302011
 Ion Ratio Lower Upper
 105 100
 120 44.5 23.3 69.8









#85
 sec-Butylbenzene
 Concen: 21.149 ug/l
 RT: 13.439 min Scan# 1964
 Delta R.T. 0.000 min
 Lab File: VN074292.D
 Acq: 02 Sep 2022 17:59




Instrument :
 MSVOA_N
 ClientSampleId :
 VN0902WBS01


Tgt Ion:105 Resp: 389044
 Ion Ratio Lower Upper
 105 100
 134 19.6 10.4 31.1



#86
 p-Isopropyltoluene
 Concen: 21.737 ug/l
 RT: 13.557 min Scan# 1984
 Delta R.T. -0.000 min
 Lab File: VN074292.D
 Acq: 02 Sep 2022 17:59

Tgt Ion:119 Resp: 311498
 Ion Ratio Lower Upper
 119 100
 134 25.8 13.3 39.8
 91 24.0 11.3 33.9









#87
1,3-Dichlorobenzene
Concen: 21.271 ug/l
RT: 13.551 min Scan# 1983
Delta R.T. -0.000 min
Lab File: VN074292.D
Acq: 02 Sep 2022 17:59




Instrument :
MSVOA_N
ClientSampleId :
VN0902WBS01


Tgt Ion:146 Resp: 156502
Ion Ratio Lower Upper
146 100
111 40.1 20.0 60.0
148 61.2 32.0 96.0



#88
1,4-Dichlorobenzene
Concen: 20.749 ug/l
RT: 13.633 min Scan# 1997
Delta R.T. 0.000 min
Lab File: VN074292.D
Acq: 02 Sep 2022 17:59

Tgt Ion:146 Resp: 151458
Ion Ratio Lower Upper
146 100
111 41.9 19.1 57.1
148 64.2 31.4 94.3





#89

n-Butylbenzene

Concen: 20.480 ug/l

RT: 13.880 min Scan# 2039

Delta R.T. 0.000 min

Lab File: VN074292.D

Acq: 02 Sep 2022 17:59

Instrument :

MSVOA_N

ClientSampleId :

VN0902WBS01





Tgt Ion: 91 Resp: 256044

Ion Ratio Lower Upper

91 100

92 52.1 26.9 80.7

134 25.5 13.6 40.8



#90

Hexachloroethane

Concen: 19.144 ug/l

RT: 14.151 min Scan# 2085

Delta R.T. -0.000 min

Lab File: VN074292.D




Acq: 02 Sep 2022 17:59


Tgt Ion: 117 Resp: 53366

Ion Ratio Lower Upper

117 100

201 67.5 40.6 122.0









#91
1,2-Dichlorobenzene
Concen: 20.062 ug/l
RT: 13.921 min Scan# 2046
Delta R.T. -0.006 min
Lab File: VN074292.D
Acq: 02 Sep 2022 17:59




Instrument :
MSVOA_N
ClientSampleId :
VN0902WBS01


Tgt Ion:146 Resp: 154724
Ion Ratio Lower Upper
146 100
111 41.7 20.9 62.8
148 62.3 32.1 96.3



#92
1,2-Dibromo-3-Chloropropane
Concen: 17.376 ug/l
RT: 14.539 min Scan# 2151
Delta R.T. 0.000 min
Lab File: VN074292.D
Acq: 02 Sep 2022 17:59

Tgt Ion: 75 Resp: 25341
Ion Ratio Lower Upper
75 100
155 84.4 44.9 134.5
157 109.7 59.0 176.8









#93
1,2,4-Trichlorobenzene
Concen: 20.089 ug/l
RT: 15.198 min Scan# 21
Delta R.T. -0.000 min
Lab File: VN074292.D
Acq: 02 Sep 2022 17:59




Instrument :
MSVOA_N
ClientSampleId :
VN0902WBS01


Tgt Ion:180 Resp: 74721
Ion Ratio Lower Upper
180 100
182 93.0 47.5 142.5
145 31.4 14.5 43.5



#94
Hexachlorobutadiene
Concen: 22.053 ug/l
RT: 15.298 min Scan# 2280
Delta R.T. -0.000 min
Lab File: VN074292.D
Acq: 02 Sep 2022 17:59

Tgt Ion:225 Resp: 35616
Ion Ratio Lower Upper
225 100
223 62.6 32.4 97.2
227 61.5 31.8 95.4





#95

Naphthalene

Concen: 19.366 ug/l

RT: 15.433 min Scan# 21

Delta R.T. 0.000 min

Lab File: VN074292.D

Acq: 02 Sep 2022 17:59

Instrument :

MSVOA_N

ClientSampleId :

VN0902WBS01





Tgt Ion:128 Resp: 282713

Ion Ratio Lower Upper

128 100

127 14.2 10.3 15.5

129 10.4 8.6 13.0



#96

1,2,3-Trichlorobenzene

Concen: 20.064 ug/l

RT: 15.627 min Scan# 2336

Delta R.T. 0.000 min

Lab File: VN074292.D

Acq: 02 Sep 2022 17:59

Tgt Ion:180 Resp: 77070

Ion Ratio Lower Upper

180 100

182 94.0 47.8 143.3

145 31.5 15.5 46.5

