

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN090222\
 Data File : VN074296.D
 Acq On : 02 Sep 2022 19:46
 Operator : JC\MD
 Sample : N4329-04DL 10X
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 MW-23ADL

Quant Time: Sep 03 04:41:18 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N090222W.M
 Quant Title : SW846 8260
 QLast Update : Sat Sep 03 04:35:33 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.016	168	272640	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.904	114	488590	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.686	117	474753	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.610	152	217144	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.369	65	186552	43.895	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	87.780%
35) Dibromofluoromethane	7.945	113	164102	50.686	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	101.380%
50) Toluene-d8	10.380	98	573550	44.925	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	89.860%
62) 4-Bromofluorobenzene	12.669	95	208426	44.922	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	89.840%
Target Compounds						
19) Methyl tert-butyl Ether	5.540	73	188052	17.679	ug/l	Qvalue 97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN090222\
Data File : VN074296.D
Acq On : 02 Sep 2022 19:46
Operator : JC\MD
Sample : N4329-04DL 10X
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 19 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
MW-23ADL

Quant Time: Sep 03 04:41:18 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N090222W.M
Quant Title : SW846 8260
QLast Update : Sat Sep 03 04:35:33 2022
Response via : Initial Calibration

