

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN090222\
 Data File : VN074303.D
 Acq On : 02 Sep 2022 22:37
 Operator : JC\MD
 Sample : N4354-09DL 20X
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 26 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 MW-27DL

Quant Time: Sep 03 04:42:26 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N090222W.M
 Quant Title : SW846 8260
 QLast Update : Sat Sep 03 04:35:33 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.022	168	245944	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.904	114	433764	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.680	117	412159	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.610	152	194355	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.369	65	168021	43.826	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	87.660%
35) Dibromofluoromethane	7.951	113	148503	51.665	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	103.340%
50) Toluene-d8	10.380	98	509654	44.966	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	89.940%
62) 4-Bromofluorobenzene	12.668	95	185757	45.086	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	90.180%
Target Compounds						
					Qvalue	
11) Tert butyl alcohol	5.281	59	13789	19.825	ug/l	95
19) Methyl tert-butyl Ether	5.545	73	15206	1.585	ug/l #	82
40) Benzene	8.392	78	315753	24.535	ug/l	96
52) Toluene	10.445	92	10693	1.345	ug/l	91
67) Ethyl Benzene	11.786	91	117088	7.050	ug/l	96
68) m/p-Xylenes	11.898	106	51283	8.271	ug/l	97
69) o-Xylene	12.221	106	32962	5.327	ug/l	100
73) Isopropylbenzene	12.515	105	5388	0.317	ug/l	89
78) n-propylbenzene	12.857	91	12317	0.633	ug/l	96
80) 1,3,5-Trimethylbenzene	13.004	105	7498	0.549	ug/l	100
84) 1,2,4-Trimethylbenzene	13.304	105	52064	3.767	ug/l	99
95) Naphthalene	15.433	128	33075	3.079	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN090222\
Data File : VN074303.D
Acq On : 02 Sep 2022 22:37
Operator : JC\MD
Sample : N4354-09DL 20X
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 26 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
MW-27DL

Quant Time: Sep 03 04:42:26 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N090222W.M
Quant Title : SW846 8260
QLast Update : Sat Sep 03 04:35:33 2022
Response via : Initial Calibration

