

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN090222\
 Data File : VN074314.D
 Acq On : 03 Sep 2022 03:04
 Operator : JC\MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 37 Sample Multiplier: 1

Instrument :
 MSVOA_N
 LabSampleId :
 VSTDCCC050

Quant Time: Sep 03 07:09:37 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N090222W.M
 Quant Title : SW846 8260
 QLast Update : Sat Sep 03 04:35:33 2022
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	1.000	1.000	0.0	102	0.00
2 T	Dichlorodifluoromethane	0.474	0.448	5.5	99	0.00
3 P	Chloromethane	0.683	0.572	16.3	88	0.00
4 C	Vinyl Chloride	1.271	1.153	9.3#	92	0.00
5 T	Bromomethane	0.855	0.514	39.9#	69	0.01
6 T	Chloroethane	1.001	1.029	-2.8	103	0.01
7 T	Trichlorofluoromethane	0.898	0.845	5.9	107	0.00
8 T	Diethyl Ether	0.366	0.347	5.2	104	0.01
9 T	1,1,2-Trichlorotrifluoroeth	0.545	0.505	7.3	105	0.01
10 T	Methyl Iodide	0.299	0.250	16.4	72	0.00
11 T	Tert butyl alcohol	0.141	0.130	7.8	97	0.00
12 CM	1,1-Dichloroethene	0.492	0.483	1.8#	107	0.00
13 T	Acrolein	0.095	0.088	7.4	106	0.00
14 T	Allyl chloride	0.791	0.738	6.7	99	0.00
15 T	Acrylonitrile	0.389	0.370	4.9	100	0.00
16 T	Acetone	0.326	0.303	7.1	103	0.00
17 T	Carbon Disulfide	1.229	1.173	4.6	99	0.01
18 T	Methyl Acetate	0.980	0.854	12.9	101	0.00
19 T	Methyl tert-butyl Ether	1.951	2.001	-2.6	105	0.00
20 T	Methylene Chloride	0.675	0.604	10.5	106	0.00
21 T	trans-1,2-Dichloroethene	0.564	0.551	2.3	101	0.00
22 T	Diisopropyl ether	1.858	1.974	-6.2	101	0.00
23 T	Vinyl Acetate	1.615	1.703	-5.4	100	0.00
24 P	1,1-Dichloroethane	1.119	1.102	1.5	103	0.00
25 T	2-Butanone	0.550	0.540	1.8	99	0.00
26 T	2,2-Dichloropropane	0.909	0.642	29.4#	74	0.00
27 T	cis-1,2-Dichloroethene	0.718	0.721	-0.4	105	0.00
28 T	Bromochloromethane	0.394	0.364	7.6	97	0.00
29 T	Tetrahydrofuran	0.351	0.358	-2.0	101	0.00
30 C	Chloroform	1.238	1.264	-2.1#	105	0.00
31 T	Cyclohexane	1.041	0.987	5.2	103	0.00
32 T	1,1,1-Trichloroethane	1.027	1.072	-4.4	104	0.00
33 S	1,2-Dichloroethane-d4	0.779	0.830	-6.5	102	0.00
34 I	1,4-Difluorobenzene	1.000	1.000	0.0	103	0.00
35 S	Dibromofluoromethane	0.331	0.345	-4.2	102	0.00
36 T	1,1-Dichloropropene	0.470	0.463	1.5	99	0.00
37 T	Ethyl Acetate	0.615	0.579	5.9	103	0.00
38 T	Carbon Tetrachloride	0.475	0.470	1.1	102	0.00
39 T	Methylcyclohexane	0.567	0.564	0.5	98	0.00
40 TM	Benzene	1.483	1.484	-0.1	101	0.00
41 T	Methacrylonitrile	0.277	0.245	11.6	98	0.00
42 TM	1,2-Dichloroethane	0.552	0.534	3.3	103	0.00
43 T	Isopropyl Acetate	0.894	0.889	0.6	99	0.00
44 TM	Trichloroethene	0.353	0.346	2.0	104	0.00
45 C	1,2-Dichloropropane	0.369	0.375	-1.6#	104	0.00
46 T	Dibromomethane	0.281	0.286	-1.8	105	0.00
47 T	Bromodichloromethane	0.516	0.534	-3.5	104	0.00
48 T	Methyl methacrylate	0.393	0.377	4.1	101	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN090222\
 Data File : VN074314.D
 Acq On : 03 Sep 2022 03:04
 Operator : JC\MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 37 Sample Multiplier: 1

Instrument :
 MSVOA_N
 LabSampleId :
 VSTDCCC050

Quant Time: Sep 03 07:09:37 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N090222W.M
 Quant Title : SW846 8260
 QLast Update : Sat Sep 03 04:35:33 2022
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
49 T	1,4-Dioxane	0.010	0.010	0.0	100	0.00
50 S	Toluene-d8	1.306	1.402	-7.4	101	0.00
51 T	4-Methyl-2-Pentanone	0.625	0.646	-3.4	100	0.00
52 CM	Toluene	0.917	0.986	-7.5#	103	0.00
53 T	t-1,3-Dichloropropene	0.525	0.539	-2.7	97	0.00
54 T	cis-1,3-Dichloropropene	0.576	0.576	0.0	98	0.00
55 T	1,1,2-Trichloroethane	0.393	0.401	-2.0	103	0.00
56 T	Ethyl methacrylate	0.586	0.643	-9.7	102	0.00
57 T	1,3-Dichloropropane	0.673	0.677	-0.6	103	0.00
58 T	2-Chloroethyl Vinyl ether	0.286	0.285	0.3	94	0.00
59 T	2-Hexanone	0.473	0.500	-5.7	99	0.00
60 T	Dibromochloromethane	0.390	0.411	-5.4	102	0.00
61 T	1,2-Dibromoethane	0.394	0.411	-4.3	103	0.00
62 S	4-Bromofluorobenzene	0.412	0.475	-15.3	104	0.00
63 I	Chlorobenzene-d5	1.000	1.000	0.0	105	0.00
64 T	Tetrachloroethene	0.314	0.307	2.2	107	0.00
65 PM	Chlorobenzene	1.098	1.088	0.9	104	0.00
66 T	1,1,1,2-Tetrachloroethane	0.406	0.400	1.5	101	0.00
67 C	Ethyl Benzene	2.015	2.085	-3.5#	104	0.00
68 T	m/p-Xylenes	0.752	0.791	-5.2	105	0.00
69 T	o-Xylene	0.751	0.780	-3.9	104	0.00
70 T	Styrene	1.202	1.307	-8.7	104	0.00
71 P	Bromoform	0.294	0.307	-4.4	103	0.00
72 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	103	0.00
73 T	Isopropylbenzene	4.378	4.392	-0.3	103	0.00
74 T	N-amyl acetate	1.867	1.871	-0.2	100	0.00
75 P	1,1,2,2-Tetrachloroethane	1.625	1.555	4.3	104	0.00
76 T	1,2,3-Trichloropropane	1.416	1.232	13.0	100	0.00
77 T	Bromobenzene	0.972	0.948	2.5	102	0.00
78 T	n-propylbenzene	5.005	5.156	-3.0	102	0.00
79 T	2-Chlorotoluene	3.073	2.954	3.9	102	0.00
80 T	1,3,5-Trimethylbenzene	3.511	3.658	-4.2	103	0.00
81 T	trans-1,4-Dichloro-2-butene	0.436	0.370	15.1	84	0.00
82 T	4-Chlorotoluene	2.885	2.907	-0.8	103	0.00
83 T	tert-Butylbenzene	3.186	3.190	-0.1	103	0.00
84 T	1,2,4-Trimethylbenzene	3.556	3.774	-6.1	105	0.00
85 T	sec-Butylbenzene	4.448	4.584	-3.1	103	0.00
86 T	p-Isopropyltoluene	3.465	3.733	-7.7	102	0.00
87 T	1,3-Dichlorobenzene	1.779	1.777	0.1	102	0.00
88 T	1,4-Dichlorobenzene	1.895	1.737	8.3	101	0.00
89 T	n-Butylbenzene	3.023	3.131	-3.6	100	0.00
90 T	Hexachloroethane	0.674	0.615	8.8	98	0.00
91 T	1,2-Dichlorobenzene	1.865	1.781	4.5	104	0.00
92 T	1,2-Dibromo-3-Chloropropane	0.353	0.320	9.3	97	0.00
93 T	1,2,4-Trichlorobenzene	0.962	0.879	8.6	101	0.00
94 T	Hexachlorobutadiene	0.422	0.364	13.7	97	0.00
95 T	Naphthalene	3.832	3.643	4.9	105	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN090222\
Data File : VN074314.D
Acq On : 03 Sep 2022 03:04
Operator : JC\MD
Sample : VSTDCCC050
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 37 Sample Multiplier: 1

Instrument :
MSVOA_N
LabSampleId :
VSTDCCC050

Quant Time: Sep 03 07:09:37 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N090222W.M
Quant Title : SW846 8260
QLast Update : Sat Sep 03 04:35:33 2022
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
96 T	1,2,3-Trichlorobenzene	1.027	0.921	10.3	103	0.00

(#) = Out of Range SPCC's out = 0 CCC's out = 6