



Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN090222\
Data File : VN074279.D
Acq On : 02 Sep 2022 10:35
Operator : JC\MD
Sample : BFB
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
BFB

Integration File: RTEINT.P

Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N090222W.M
Title : SW846 8260
Last Update : Sat Sep 03 04:35:33 2022



AutoFind: Scans 1832, 1833, 1834; Background Corrected with Scan 1823

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	23.8	11861	PASS
75	95	30	60	55.7	27715	PASS
95	95	100	100	100.0	49779	PASS
96	95	5	9	5.8	2897	PASS
173	174	0.00	2	0.6	232	PASS
174	95	50	100	77.5	38571	PASS
175	174	5	9	7.9	3044	PASS
176	174	95	101	95.4	36784	PASS
177	176	5	9	6.5	2375	PASS