

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN090319\
 Data File : VN057669.D
 Acq On : 3 Sep 2019 18:13
 Operator : JC/SP
 Sample : K4651-01
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 TT-TB-20190830

Quant Time: Sep 04 03:07:45 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA N\METHODS\624N090319W.M
 Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
 QLast Update : Wed Sep 04 02:08:50 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Bromochloromethane	7.18	128	57563	30.00	ug/l	0.00
28) 1,4-Difluorobenzene	8.58	114	344562	30.00	ug/l	0.00
57) Chlorobenzene-d5	11.41	117	261757	30.00	ug/l	0.00

System Monitoring Compounds						
27) 1,2-Dichloroethane-d4	8.02	65	187213	30.59	ug/l	0.00
Spiked Amount	30.000	Range	50 - 169	Recovery	=	101.97%
60) 4-Bromofluorobenzene	12.40	95	101971	22.80	ug/l	0.00
Spiked Amount	30.000	Range	56 - 143	Recovery	=	76.00%
63) Toluene-d8	10.09	98	419602	33.29	ug/l	0.00
Spiked Amount	30.000	Range	66 - 137	Recovery	=	110.97%

Target Compounds					Ovalue
3) Chloromethane	2.05	50	4324	0.989 ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN090319\
Data File : VN057669.D
Acq On : 3 Sep 2019 18:13
Operator : JC/SP
Sample : K4651-01
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
TT-TB-20190830

Quant Time: Sep 04 03:07:45 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA N\METHODS\624N090319W.M
Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
QLast Update : Wed Sep 04 02:08:50 2019
Response via : Initial Calibration

