

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN090319\
 Data File : VN057670.D
 Acq On : 3 Sep 2019 18:35
 Operator : JC/SP
 Sample : K4651-01
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 TT-TB-20190830

Quant Time: Sep 04 03:10:05 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA N\METHODS\624N090319W.M
 Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
 QLast Update : Wed Sep 04 02:08:50 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Bromochloromethane	7.18	128	58248	30.00	ug/l	0.00
28) 1,4-Difluorobenzene	8.58	114	356118	30.00	ug/l	0.00
57) Chlorobenzene-d5	11.41	117	279315	30.00	ug/l	0.00

System Monitoring Compounds						
27) 1,2-Dichloroethane-d4	8.01	65	191562	30.94	ug/l	0.00
Spiked Amount	30.000	Range	50 - 169	Recovery	=	103.13%
60) 4-Bromofluorobenzene	12.40	95	106564	22.33	ug/l	0.00
Spiked Amount	30.000	Range	56 - 143	Recovery	=	74.43%
63) Toluene-d8	10.09	98	435358	32.37	ug/l	0.00
Spiked Amount	30.000	Range	66 - 137	Recovery	=	107.90%

Target Compounds	Ovalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN090319\
Data File : VN057670.D
Acq On : 3 Sep 2019 18:35
Operator : JC/SP
Sample : K4651-01
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
TT-TB-20190830

Quant Time: Sep 04 03:10:05 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA N\METHODS\624N090319W.M
Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
QLast Update : Wed Sep 04 02:08:50 2019
Response via : Initial Calibration

