

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN090319\
 Data File : VN057672.D
 Acq On : 3 Sep 2019 19:20
 Operator : JC/SP
 Sample : K4651-02
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 TT-006-IDW/GW-20190830

Quant Time: Sep 05 15:24:32 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA N\METHODS\624N090319W.M
 Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
 QLast Update : Wed Sep 04 02:08:50 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Bromochloromethane	7.18	128	57649	30.00	ug/l	0.00
28) 1,4-Difluorobenzene	8.58	114	347496	30.00	ug/l	0.00
57) Chlorobenzene-d5	11.41	117	263401	30.00	ug/l	0.00

System Monitoring Compounds						
27) 1,2-Dichloroethane-d4	8.01	65	185118	30.21	ug/l	0.00
Spiked Amount	30.000	Range	50 - 169	Recovery	=	100.70%
60) 4-Bromofluorobenzene	12.40	95	103345	22.97	ug/l	0.00
Spiked Amount	30.000	Range	56 - 143	Recovery	=	76.57%
63) Toluene-d8	10.09	98	423772	33.41	ug/l	0.00
Spiked Amount	30.000	Range	66 - 137	Recovery	=	111.37%

Target Compounds					Ovalue
37) Trichloroethene	8.83	130	34192	9.034 ug/l	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN090319\
Data File : VN057672.D
Acq On : 3 Sep 2019 19:20
Operator : JC/SP
Sample : K4651-02
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
TT-006-IDW/GW-20190830

Quant Time: Sep 05 15:24:32 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA N\METHODS\624N090319W.M
Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
QLast Update : Wed Sep 04 02:08:50 2019
Response via : Initial Calibration

