

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN090320\
 Data File : VN063217.D
 Acq On : 3 Sep 2020 14:00
 Operator : JC/MD
 Sample : PB131386ZHE#06
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB131386ZHE#06

Manual Integrations
 APPROVED

MMDadoda
 9/4/2020 11:14:19 AM

Quant Time: Sep 04 04:34:50 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N083120W.M
 Quant Title : SW846 8260
 QLast Update : Tue Sep 01 04:21:28 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.63	168	285379	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.55	114	562882	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.38	117	575392	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.32	152	224762	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.99	65	229745	49.25	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.50%	
35) Dibromofluoromethane	7.55	113	178704	49.49	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.98%	
50) Toluene-d8	10.06	98	724626	52.13	ug/l	0.00
Spiked Amount 50.000			Recovery	=	104.26%	
62) 4-Bromofluorobenzene	12.38	95	250618	51.55	ug/l	0.00
Spiked Amount 50.000			Recovery	=	103.10%	
Target Compounds						
16) Acetone	3.78	43	28037	21.142	ug/l	97
18) Methyl Acetate	4.30	43	7203m	2.118	ug/l	
20) Methylene Chloride	4.50	84	15822	3.788	ug/l	89
43) Isopropyl Acetate	8.13	43	101995	11.558	ug/l #	91

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN090320\
Data File : VN063217.D
Acq On : 3 Sep 2020 14:00
Operator : JC/MD
Sample : PB131386ZHE#06
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB131386ZHE#06

Manual Integrations
APPROVED

MMDadoda
9/4/2020 11:14:19 AM

Quant Time: Sep 04 04:34:50 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N083120W.M
Quant Title : SW846 8260
QLast Update : Tue Sep 01 04:21:28 2020
Response via : Initial Calibration

