

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN090320\
 Data File : VN063218.D
 Acq On : 3 Sep 2020 14:24
 Operator : JC/MD
 Sample : PB131386ZHE#07
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB131386ZHE#07

Manual Integrations
 APPROVED

MMDadoda
 9/4/2020 11:14:14 AM

Quant Time: Sep 04 04:37:51 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N083120W.M
 Quant Title : SW846 8260
 QLast Update : Tue Sep 01 04:21:28 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.63	168	291082	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.55	114	592181	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.38	117	603028	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.32	152	228258	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.99	65	240815	50.61	ug/l	0.00
Spiked Amount 50.000			Recovery	=	101.22%	
35) Dibromofluoromethane	7.55	113	187237	49.28	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.56%	
50) Toluene-d8	10.06	98	765177	52.33	ug/l	0.00
Spiked Amount 50.000			Recovery	=	104.66%	
62) 4-Bromofluorobenzene	12.38	95	266522	52.10	ug/l	0.00
Spiked Amount 50.000			Recovery	=	104.20%	
Target Compounds						
16) Acetone	3.78	43	30458	22.517	ug/l	91
18) Methyl Acetate	4.29	43	5280m	1.522	ug/l	
20) Methylene Chloride	4.50	84	15927m	3.732	ug/l	
43) Isopropyl Acetate	8.13	43	103663	11.165	ug/l #	89

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN090320\
Data File : VN063218.D
Acq On : 3 Sep 2020 14:24
Operator : JC/MD
Sample : PB131386ZHE#07
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB131386ZHE#07

Manual Integrations
APPROVED

MMDadoda
9/4/2020 11:14:14 AM

Quant Time: Sep 04 04:37:51 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N083120W.M
Quant Title : SW846 8260
QLast Update : Tue Sep 01 04:21:28 2020
Response via : Initial Calibration

