

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN090320\
 Data File : VN063219.D
 Acq On : 3 Sep 2020 14:48
 Operator : JC/MD
 Sample : PB131386ZHE#08
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB131386ZHE#08

Manual Integrations
 APPROVED

MMDadoda
 9/4/2020 11:14:08 AM

Quant Time: Sep 04 04:40:23 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N083120W.M
 Quant Title : SW846 8260
 QLast Update : Tue Sep 01 04:21:28 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.62	168	298984	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.55	114	594229	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.38	117	595979	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.32	152	230196	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.99	65	238961	48.89	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.78%	
35) Dibromofluoromethane	7.55	113	184143	48.30	ug/l	0.00
Spiked Amount 50.000			Recovery	=	96.60%	
50) Toluene-d8	10.06	98	750016	51.11	ug/l	0.00
Spiked Amount 50.000			Recovery	=	102.22%	
62) 4-Bromofluorobenzene	12.38	95	264107	51.45	ug/l	0.00
Spiked Amount 50.000			Recovery	=	102.90%	
Target Compounds						
16) Acetone	3.78	43	27667	19.913	ug/l	90
18) Methyl Acetate	4.30	43	6098m	1.712	ug/l	
20) Methylene Chloride	4.51	84	19271	4.485	ug/l	94
43) Isopropyl Acetate	8.13	43	100133	10.748	ug/l #	90

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN090320\
Data File : VN063219.D
Acq On : 3 Sep 2020 14:48
Operator : JC/MD
Sample : PB131386ZHE#08
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB131386ZHE#08

Manual Integrations
APPROVED

MMDadoda
9/4/2020 11:14:08 AM

Quant Time: Sep 04 04:40:23 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N083120W.M
Quant Title : SW846 8260
QLast Update : Tue Sep 01 04:21:28 2020
Response via : Initial Calibration

