

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN090320\
 Data File : VN063220.D
 Acq On : 3 Sep 2020 15:12
 Operator : JC/MD
 Sample : PB131386ZHE#09
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB131386ZHE#09

Manual Integrations
 APPROVED

MMDadoda
 9/4/2020 11:14:03 AM

Quant Time: Sep 04 04:42:51 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N083120W.M
 Quant Title : SW846 8260
 QLast Update : Tue Sep 01 04:21:28 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.63	168	294010	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.55	114	580014	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.38	117	597354	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.32	152	230112	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.99	65	235816	49.06	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.12%	
35) Dibromofluoromethane	7.55	113	184028	49.46	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.92%	
50) Toluene-d8	10.06	98	753447	52.61	ug/l	0.00
Spiked Amount 50.000			Recovery	=	105.22%	
62) 4-Bromofluorobenzene	12.38	95	267494	53.39	ug/l	0.00
Spiked Amount 50.000			Recovery	=	106.78%	
Target Compounds						
16) Acetone	3.77	43	30349	22.213	ug/l	97
18) Methyl Acetate	4.29	43	6078m	1.735	ug/l	
20) Methylene Chloride	4.51	84	25442	6.192	ug/l	93
43) Isopropyl Acetate	8.13	43	101063	11.114	ug/l #	90

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN090320\
Data File : VN063220.D
Acq On : 3 Sep 2020 15:12
Operator : JC/MD
Sample : PB131386ZHE#09
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB131386ZHE#09

Manual Integrations
APPROVED

MMDadoda
9/4/2020 11:14:03 AM

Quant Time: Sep 04 04:42:51 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N083120W.M
Quant Title : SW846 8260
QLast Update : Tue Sep 01 04:21:28 2020
Response via : Initial Calibration

