

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN090320\
 Data File : VN063221.D
 Acq On : 3 Sep 2020 15:36
 Operator : JC/MD
 Sample : PB131386ZHE#10
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB131386ZHE#10

Quant Time: Sep 04 04:44:56 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N083120W.M
 Quant Title : SW846 8260
 QLast Update : Tue Sep 01 04:21:28 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.63	168	288619	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.55	114	582209	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.38	117	592446	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.32	152	227755	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.99	65	235810	49.98	ug/l	0.00
Spiked Amount 50.000			Recovery	=	99.96%	
35) Dibromofluoromethane	7.55	113	183197	49.05	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.10%	
50) Toluene-d8	10.06	98	753418	52.41	ug/l	0.00
Spiked Amount 50.000			Recovery	=	104.82%	
62) 4-Bromofluorobenzene	12.38	95	255842	50.87	ug/l	0.00
Spiked Amount 50.000			Recovery	=	101.74%	
Target Compounds						
16) Acetone	3.78	43	41357	30.836	ug/l	97
18) Methyl Acetate	4.28	43	5188	1.509	ug/l #	77
20) Methylene Chloride	4.51	84	16870	4.021	ug/l	92
43) Isopropyl Acetate	8.13	43	97794	10.714	ug/l #	88

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN090320\
Data File : VN063221.D
Acq On : 3 Sep 2020 15:36
Operator : JC/MD
Sample : PB131386ZHE#10
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB131386ZHE#10

Quant Time: Sep 04 04:44:56 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N083120W.M
Quant Title : SW846 8260
QLast Update : Tue Sep 01 04:21:28 2020
Response via : Initial Calibration

