

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN090320\
 Data File : VN063228.D
 Acq On : 3 Sep 2020 18:24
 Operator : JC/MD
 Sample : L3902-06
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 SA-MW131I-20200901

Quant Time: Sep 04 05:25:08 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N083120W.M
 Quant Title : SW846 8260
 QLast Update : Tue Sep 01 04:21:28 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.63	168	286718	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.55	114	564137	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.38	117	552731	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.32	152	210342	50.00	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.99	65	227619	48.56	ug/l	0.00
Spiked Amount	50.000		Recovery	=	97.12%	
35) Dibromofluoromethane	7.55	113	178004	49.18	ug/l	0.00
Spiked Amount	50.000		Recovery	=	98.36%	
50) Toluene-d8	10.06	98	705281	50.63	ug/l	0.00
Spiked Amount	50.000		Recovery	=	101.26%	
62) 4-Bromofluorobenzene	12.38	95	244011	50.07	ug/l	0.00
Spiked Amount	50.000		Recovery	=	100.14%	

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN090320\
Data File : VN063228.D
Acq On : 3 Sep 2020 18:24
Operator : JC/MD
Sample : L3902-06
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 22 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
SA-MW131I-20200901

Quant Time: Sep 04 05:25:08 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N083120W.M
Quant Title : SW846 8260
QLast Update : Tue Sep 01 04:21:28 2020
Response via : Initial Calibration

