

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN090321\
 Data File : VN068413.D
 Acq On : 3 Sep 2021 17:30
 Operator : JC/MD
 Sample : M3644-04RE
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 191-0902-DUPRE

Manual Integrations
 APPROVED

MMDadoda
 9/7/2021 4:39:21 PM

Quant Time: Sep 04 03:13:02 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\624N090321W.M
 Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
 QLast Update : Fri Sep 03 13:24:01 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Bromochloromethane	7.667	128	88474	30.000	ug/l	0.00
28) 1,4-Difluorobenzene	8.973	114	456728	30.000	ug/l	# 0.00
57) Chlorobenzene-d5	11.746	117	393697	30.000	ug/l	# 0.00
System Monitoring Compounds						
27) 1,2-Dichloroethane-d4	8.440	65	176904	33.726	ug/l	0.00
Spiked Amount	30.000	Range	91 - 110	Recovery	=	112.433%#
60) 4-Bromofluorobenzene	12.736	95	148587	26.651	ug/l	0.00
Spiked Amount	30.000	Range	63 - 112	Recovery	=	88.833%
63) Toluene-d8	10.446	98	544497	31.057	ug/l	0.00
Spiked Amount	30.000	Range	91 - 112	Recovery	=	103.533%
Target Compounds						Qvalue
25) Chloroform	7.825	83	3305m	0.360	ug/l	

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN090321\
Data File : VN068413.D
Acq On : 3 Sep 2021 17:30
Operator : JC/MD
Sample : M3644-04RE
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
191-0902-DUPRE

Manual Integrations
APPROVED

MMDadoda
9/7/2021 4:39:21 PM

Quant Time: Sep 04 03:13:02 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\624N090321W.M
Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
QLast Update : Fri Sep 03 13:24:01 2021
Response via : Initial Calibration

