

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN090321\
 Data File : VN068414.D
 Acq On : 3 Sep 2021 17:52
 Operator : JC/MD
 Sample : M3644-05RE
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 TRIP-BLANKRE

Quant Time: Sep 04 03:13:15 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\624N090321W.M
 Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
 QLast Update : Fri Sep 03 13:24:01 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Bromochloromethane	7.665	128	79289	30.000	ug/l	0.00
28) 1,4-Difluorobenzene	8.971	114	415938	30.000	ug/l	# 0.00
57) Chlorobenzene-d5	11.749	117	352878	30.000	ug/l	# 0.00

System Monitoring Compounds						
27) 1,2-Dichloroethane-d4	8.440	65	160225	34.085	ug/l	0.00
Spiked Amount	30.000	Range	91 - 110	Recovery	=	113.633%#
60) 4-Bromofluorobenzene	12.734	95	132894	26.594	ug/l	0.00
Spiked Amount	30.000	Range	63 - 112	Recovery	=	88.633%
63) Toluene-d8	10.443	98	496161	31.573	ug/l	0.00
Spiked Amount	30.000	Range	91 - 112	Recovery	=	105.233%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN090321\
Data File : VN068414.D
Acq On : 3 Sep 2021 17:52
Operator : JC/MD
Sample : M3644-05RE
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
TRIP-BLANKRE

Quant Time: Sep 04 03:13:15 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\624N090321W.M
Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
QLast Update : Fri Sep 03 13:24:01 2021
Response via : Initial Calibration

