

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN090322\
 Data File : VN074334.D
 Acq On : 03 Sep 2022 22:13
 Operator : JC\MD
 Sample : N4454-01DL 5X
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 MW-13A-8-13-083022DL

Quant Time: Sep 05 05:41:51 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N090222W.M
 Quant Title : SW846 8260
 QLast Update : Sat Sep 03 04:35:33 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.016	168	333055	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.904	114	609909	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.680	117	591410	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.615	152	270002	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.375	65	238986	46.032	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	92.060%
35) Dibromofluoromethane	7.957	113	208361	51.555	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	103.100%
50) Toluene-d8	10.380	98	725755	45.540	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	91.080%
62) 4-Bromofluorobenzene	12.668	95	261858	45.194	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	90.380%
Target Compounds						
					Qvalue	
4) Vinyl Chloride	2.404	62	59465	7.023	ug/l #	63
27) cis-1,2-Dichloroethene	7.251	96	114077	23.865	ug/l	87
44) Trichloroethene	9.151	130	38324	8.899	ug/l	93

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN090322\
Data File : VN074334.D
Acq On : 03 Sep 2022 22:13
Operator : JC\MD
Sample : N4454-01DL 5X
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 20 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
MW-13A-8-13-083022DL

Quant Time: Sep 05 05:41:51 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N090222W.M
Quant Title : SW846 8260
QLast Update : Sat Sep 03 04:35:33 2022
Response via : Initial Calibration

