

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN090322\
 Data File : VN074338.D
 Acq On : 03 Sep 2022 23:50
 Operator : JC\MD
 Sample : N4355-11
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 24 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 MW-36

Quant Time: Sep 05 05:43:10 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N090222W.M
 Quant Title : SW846 8260
 QLast Update : Sat Sep 03 04:35:33 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.016	168	313763	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	8.904	114	557656	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.686	117	547783	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.610	152	249072	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.369	65	229280	46.878	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	93.760%
35) Dibromofluoromethane	7.951	113	200728	54.320	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	108.640%
50) Toluene-d8	10.380	98	680496	46.701	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	93.400%
62) 4-Bromofluorobenzene	12.668	95	247319	46.590	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	93.180%
Target Compounds						
16) Acetone	4.234	43	7086	3.467	ug/l	# 80

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN090322\
Data File : VN074338.D
Acq On : 03 Sep 2022 23:50
Operator : JC\MD
Sample : N4355-11
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 24 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
MW-36

Quant Time: Sep 05 05:43:10 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N090222W.M
Quant Title : SW846 8260
QLast Update : Sat Sep 03 04:35:33 2022
Response via : Initial Calibration

