

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN090324\
 Data File : VN083615.D
 Acq On : 03 Sep 2024 10:17
 Operator : JC\MD
 Sample : VN0903MBL01
 Misc : 5.00g/10mL/100uL/5.00mL/MSVOA_N/MEOH
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VN0903MBL01

Quant Time: Sep 04 02:53:38 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N083024W.M
 Quant Title : SW846 8260
 QLast Update : Sat Aug 31 00:48:10 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.224	168	139170	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.100	114	276121	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	281927	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	125479	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.577	65	99370	51.089	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	102.180%
35) Dibromofluoromethane	8.159	113	80967	48.719	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	97.440%
50) Toluene-d8	10.565	98	293816	45.454	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	90.900%
62) 4-Bromofluorobenzene	12.847	95	124476	51.921	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	=	103.840%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN090324\
Data File : VN083615.D
Acq On : 03 Sep 2024 10:17
Operator : JC\MD
Sample : VN0903MBL01
Misc : 5.00g/10mL/100uL/5.00mL/MSVOA_N/MEOH
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VN0903MBL01

Quant Time: Sep 04 02:53:38 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N083024W.M
Quant Title : SW846 8260
QLast Update : Sat Aug 31 00:48:10 2024
Response via : Initial Calibration

