

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN090324\
 Data File : VN083619.D
 Acq On : 03 Sep 2024 12:45
 Operator : JC\MD
 Sample : P3787-02
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 SB-1

Quant Time: Sep 03 13:05:39 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N083024W.M
 Quant Title : SW846 8260
 QLast Update : Sat Aug 31 00:48:10 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.224	168	145628	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.100	114	301002	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	321506	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.788	152	162576	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.577	65	107946	53.037	ug/l	0.00
Spiked Amount	50.000	Range 74 - 125	Recovery	= 106.080%		
35) Dibromofluoromethane	8.159	113	87356	48.219	ug/l	0.00
Spiked Amount	50.000	Range 75 - 124	Recovery	= 96.440%		
50) Toluene-d8	10.565	98	325962	46.239	ug/l	0.00
Spiked Amount	50.000	Range 86 - 113	Recovery	= 92.480%		
62) 4-Bromofluorobenzene	12.847	95	147919	56.599	ug/l	0.00
Spiked Amount	50.000	Range 77 - 121	Recovery	= 113.200%		
Target Compounds						
					Qvalue	
16) Acetone	4.436	43	33643	41.451	ug/l	97
18) Methyl Acetate	5.018	43	4280	1.983	ug/l #	84
20) Methylene Chloride	5.265	84	6856	3.825	ug/l	87
25) 2-Butanone	7.483	43	9711	8.788	ug/l	96
40) Benzene	8.600	78	102210	12.024	ug/l	99
43) Isopropyl Acetate	8.689	43	362272	66.021	ug/l #	72
52) Toluene	10.630	92	99317	18.681	ug/l	97
67) Ethyl Benzene	11.959	91	24244	1.884	ug/l	99
68) m/p-Xylenes	12.065	106	47883	9.876	ug/l	98
69) o-Xylene	12.394	106	42532	9.076	ug/l	99
80) 1,3,5-Trimethylbenzene	13.171	105	20380	1.839	ug/l	97
84) 1,2,4-Trimethylbenzene	13.482	105	58947	5.283	ug/l	84
93) 1,2,4-Trichlorobenzene	15.394	180	3965	1.212	ug/l	94
95) Naphthalene	15.641	128	1864580	177.333	ug/l	100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN090324\
Data File : VN083619.D
Acq On : 03 Sep 2024 12:45
Operator : JC\MD
Sample : P3787-02
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
SB-1

Quant Time: Sep 03 13:05:39 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N083024W.M
Quant Title : SW846 8260
QLast Update : Sat Aug 31 00:48:10 2024
Response via : Initial Calibration

