

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN090324\
 Data File : VN083628.D
 Acq On : 03 Sep 2024 16:45
 Operator : JC\MD
 Sample : PB163072ZHE#03
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB163072ZHE#03

Quant Time: Sep 04 02:59:01 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N083024W.M
 Quant Title : SW846 8260
 QLast Update : Sat Aug 31 00:48:10 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.224	168	120720	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.100	114	258677	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	294391	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	147538	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.577	65	114410	67.811	ug/l	0.00
Spiked Amount 50.000	Range 74 - 125		Recovery = 135.620%#			
35) Dibromofluoromethane	8.165	113	92046	59.121	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery = 118.240%			
50) Toluene-d8	10.565	98	340222	55.914	ug/l	0.00
Spiked Amount 50.000	Range 86 - 113		Recovery = 111.820%			
62) 4-Bromofluorobenzene	12.847	95	166290	74.039	ug/l	0.00
Spiked Amount 50.000	Range 77 - 121		Recovery = 148.080%#			
Target Compounds					Qvalue	
16) Acetone	4.436	43	25337	37.658	ug/l	98
18) Methyl Acetate	5.018	43	10382	5.802	ug/l	97
20) Methylene Chloride	5.259	84	5255	3.537	ug/l	88
37) Ethyl Acetate	7.553	43	4256	1.801	ug/l #	89
43) Isopropyl Acetate	8.688	43	307005	65.087	ug/l #	77
95) Naphthalene	15.641	128	12854	1.347	ug/l #	94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN090324\
Data File : VN083628.D
Acq On : 03 Sep 2024 16:45
Operator : JC\MD
Sample : PB163072ZHE#03
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB163072ZHE#03

Quant Time: Sep 04 02:59:01 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N083024W.M
Quant Title : SW846 8260
QLast Update : Sat Aug 31 00:48:10 2024
Response via : Initial Calibration

