

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN090324\
 Data File : VN083629.D
 Acq On : 03 Sep 2024 17:09
 Operator : JC\MD
 Sample : P3799-07RE
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 WELL-1RE

Quant Time: Sep 04 02:59:30 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N083024W.M
 Quant Title : SW846 8260
 QLast Update : Sat Aug 31 00:48:10 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.224	168	153716	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.100	114	314981	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	338360	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	159986	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.577	65	117022	54.471	ug/l	0.00
Spiked Amount 50.000	Range 74 - 125		Recovery = 108.940%			
35) Dibromofluoromethane	8.165	113	93152	49.136	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery = 98.280%			
50) Toluene-d8	10.565	98	342632	46.441	ug/l	0.00
Spiked Amount 50.000	Range 86 - 113		Recovery = 92.880%			
62) 4-Bromofluorobenzene	12.847	95	160130	58.552	ug/l	0.00
Spiked Amount 50.000	Range 77 - 121		Recovery = 117.100%			
Target Compounds						Qvalue
18) Methyl Acetate	5.024	43	23171	10.170	ug/l	98
70) Styrene	12.412	104	7330	0.884	ug/l	91

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN083629\
Data File : VN083629.D
Acq On : 03 Sep 2024 17:09
Operator : JC\MD
Sample : P3799-07RE
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 18 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
WELL-1RE

Quant Time: Sep 04 02:59:30 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N083024W.M
Quant Title : SW846 8260
QLast Update : Sat Aug 31 00:48:10 2024
Response via : Initial Calibration

