

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN090324\
 Data File : VN083630.D
 Acq On : 03 Sep 2024 17:33
 Operator : JC\MD
 Sample : P3799-08RE
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 WELL-2RE

Quant Time: Sep 04 02:59:55 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N083024W.M
 Quant Title : SW846 8260
 QLast Update : Sat Aug 31 00:48:10 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.224	168	149272	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.100	114	312197	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	330124	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.788	152	162965	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.577	65	114774	55.015	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	110.040%
35) Dibromofluoromethane	8.165	113	91720	48.812	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	97.620%
50) Toluene-d8	10.565	98	329347	45.073	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	90.140%
62) 4-Bromofluorobenzene	12.847	95	158758	58.568	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	=	117.140%
Target Compounds						
						Qvalue
16) Acetone	4.436	43	3972	4.774	ug/l	90
18) Methyl Acetate	5.024	43	12428	5.617	ug/l	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN090324\
Data File : VN083630.D
Acq On : 03 Sep 2024 17:33
Operator : JC\MD
Sample : P3799-08RE
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 19 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
WELL-2RE

Quant Time: Sep 04 02:59:55 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N083024W.M
Quant Title : SW846 8260
QLast Update : Sat Aug 31 00:48:10 2024
Response via : Initial Calibration

