

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN090324\
 Data File : VN083631.D
 Acq On : 03 Sep 2024 17:57
 Operator : JC\MD
 Sample : P3799-09RE
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 WELL-4RE

Quant Time: Sep 04 03:00:19 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N083024W.M
 Quant Title : SW846 8260
 QLast Update : Sat Aug 31 00:48:10 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.224	168	147909	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.100	114	301724	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	322062	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.788	152	158389	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.577	65	110010	53.218	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	106.440%
35) Dibromofluoromethane	8.165	113	89966	49.540	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	99.080%
50) Toluene-d8	10.565	98	324576	45.940	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	91.880%
62) 4-Bromofluorobenzene	12.847	95	156165	59.611	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	=	119.220%
Target Compounds						
					Qvalue	
16) Acetone	4.436	43	3687	4.473	ug/l #	75
18) Methyl Acetate	5.024	43	6049	2.759	ug/l	91

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN090324\
 Data File : VN083631.D
 Acq On : 03 Sep 2024 17:57
 Operator : JC\MD
 Sample : P3799-09RE
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 MSVOA_N
ClientSampleId :
 WELL-4RE

Quant Time: Sep 04 03:00:19 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N083024W.M
 Quant Title : SW846 8260
 QLast Update : Sat Aug 31 00:48:10 2024
 Response via : Initial Calibration

