

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN090324\
 Data File : VN083634.D
 Acq On : 03 Sep 2024 19:09
 Operator : JC\MD
 Sample : P3769-11DL 2X
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 23 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 MW-13A-10.5-082724DL

Quant Time: Sep 04 05:50:06 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N083024W.M
 Quant Title : SW846 8260
 QLast Update : Sat Aug 31 00:48:10 2024
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.224	168	147056	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.100	114	300430	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	317087	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	153616	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.577	65	112952	54.958	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	= 109.920%	
35) Dibromofluoromethane	8.165	113	90150	49.856	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 99.720%	
50) Toluene-d8	10.565	98	326960	46.463	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	= 92.920%	
62) 4-Bromofluorobenzene	12.847	95	154943	59.399	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	= 118.800%	
Target Compounds						
					Qvalue	
4) Vinyl Chloride	2.512	62	36980	20.381	ug/l #	87
18) Methyl Acetate	5.030	43	13761	6.313	ug/l	98
27) cis-1,2-Dichloroethene	7.483	96	110522	55.193	ug/l	90
44) Trichloroethene	9.347	130	33187	16.260	ug/l	94
65) Chlorobenzene	11.888	112	5341	0.751	ug/l	94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN083634\
Data File : VN083634.D
Acq On : 03 Sep 2024 19:09
Operator : JC\MD
Sample : P3769-11DL 2X
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 23 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
MW-13A-10.5-082724DL

Quant Time: Sep 04 05:50:06 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N083024W.M
Quant Title : SW846 8260
QLast Update : Sat Aug 31 00:48:10 2024
Response via : Initial Calibration

