

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN090418\
 Data File : VN051005.D
 Acq On : 4 Sep 2018 9:13
 Operator : MD\SY
 Sample : VN0904WBL01
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VN0904WBL01

Quant Time: Sep 05 01:31:22 2018
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N082118W.M
 Quant Title : SW846 8260
 QLast Update : Wed Aug 22 05:38:12 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.67	168	605770	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	941340	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	781693	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.35	152	240365	50.00	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.03	65	377362	51.75	ug/l	0.00
Spiked Amount	50.000		Recovery	=	103.50%	
35) Dibromofluoromethane	7.59	113	367028	49.01	ug/l	0.00
Spiked Amount	50.000		Recovery	=	98.02%	
50) Toluene-d8	10.09	98	1291189	46.12	ug/l	0.00
Spiked Amount	50.000		Recovery	=	92.24%	
62) 4-Bromofluorobenzene	12.40	95	328308	35.29	ug/l	0.00
Spiked Amount	50.000		Recovery	=	70.58%	

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN090418\
Data File : VN051005.D
Acq On : 4 Sep 2018 9:13
Operator : MD\SY
Sample : VN0904WBL01
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VN0904WBL01

Quant Time: Sep 05 01:31:22 2018
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N082118W.M
Quant Title : SW846 8260
QLast Update : Wed Aug 22 05:38:12 2018
Response via : Initial Calibration

