

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN090418\
 Data File : VN051011.D
 Acq On : 4 Sep 2018 11:49
 Operator : MD\SY
 Sample : J4763-02DL 10X
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 002M-35TH-AVE(AUG)DL

Quant Time: Sep 04 12:18:01 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA N\METHODS\624N082818W.M
 Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
 QLast Update : Wed Aug 29 01:38:44 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Bromochloromethane	7.20	128	110338	30.00	ug/l	0.00
28) 1,4-Difluorobenzene	8.59	114	546197	30.00	ug/l	0.00
57) Chlorobenzene-d5	11.41	117	465431	30.00	ug/l	0.00

System Monitoring Compounds						
27) 1,2-Dichloroethane-d4	8.03	65	209924	29.81	ug/l	0.00
Spiked Amount	30.000	Range	50 - 169	Recovery	=	99.37%
60) 4-Bromofluorobenzene	12.40	95	149344	21.58	ug/l	0.00
Spiked Amount	30.000	Range	56 - 143	Recovery	=	71.93%
63) Toluene-d8	10.09	98	650918	29.99	ug/l	0.00
Spiked Amount	30.000	Range	66 - 137	Recovery	=	99.97%

Target Compounds					Ovalue
15) Acetone	3.82	58	131220	219.99	ug/l 95
62) Toluene	10.16	91	1185403	45.65	ug/l 100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN090418\
Data File : VN051011.D
Acq On : 4 Sep 2018 11:49
Operator : MD\SY
Sample : J4763-02DL 10X
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
002M-35TH-AVE(AUG)DL

Quant Time: Sep 04 12:18:01 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA N\METHODS\624N082818W.M
Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
QLast Update : Wed Aug 29 01:38:44 2018
Response via : Initial Calibration

