

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN090418\
 Data File : VN051024.D
 Acq On : 4 Sep 2018 17:10
 Operator : MD\SY
 Sample : J4523-04
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 23 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 WC-2

Quant Time: Sep 05 02:29:14 2018
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N082118W.M
 Quant Title : SW846 8260
 QLast Update : Wed Aug 22 05:38:12 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.67	168	569495	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	920278	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	783575	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.35	152	257759	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	8.03	65	369207	53.85	ug/l	0.00
Spiked Amount				50.000		
			Recovery	=	107.70%	
35) Dibromofluoromethane	7.60	113	356421	48.68	ug/l	0.00
Spiked Amount				50.000		
			Recovery	=	97.36%	
50) Toluene-d8	10.09	98	1243675	45.44	ug/l	0.00
Spiked Amount				50.000		
			Recovery	=	90.88%	
62) 4-Bromofluorobenzene	12.40	95	337299	37.08	ug/l	0.00
Spiked Amount				50.000		
			Recovery	=	74.16%	

Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
16) Acetone	3.83	43	13134	1.06	ug/l	99
18) Methyl Acetate	4.34	43	12668	1.96	ug/l	95
20) Methylene Chloride	4.55	84	101433	13.13	ug/l	98
30) Chloroform	7.38	83	14990	1.19	ug/l	99
43) Isopropyl Acetate	8.17	43	98065	9.14	ug/l	94
95) Naphthalene	15.13	128	13509	6.57	ug/l	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN090418\
Data File : VN051024.D
Acq On : 4 Sep 2018 17:10
Operator : MD\SY
Sample : J4523-04
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 23 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
WC-2

Quant Time: Sep 05 02:29:14 2018
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N082118W.M
Quant Title : SW846 8260
QLast Update : Wed Aug 22 05:38:12 2018
Response via : Initial Calibration

