

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN090418\
 Data File : VN051020.D
 Acq On : 4 Sep 2018 15:31
 Operator : MD\SY
 Sample : PB112553ZHE#06
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB112553ZHE#06

Quant Time: Sep 05 02:14:48 2018
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N082118W.M
 Quant Title : SW846 8260
 QLast Update : Wed Aug 22 05:38:12 2018
 Response via : Initial Calibration


Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.67	168	588429	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	956222	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	785933	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.35	152	231207	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.03	65	379694	53.60	ug/l	0.00
Spiked Amount 50.000			Recovery	=	107.20%	
35) Dibromofluoromethane	7.59	113	363057	47.72	ug/l	0.00
Spiked Amount 50.000			Recovery	=	95.44%	
50) Toluene-d8	10.09	98	1295106	45.54	ug/l	0.00
Spiked Amount 50.000			Recovery	=	91.08%	
62) 4-Bromofluorobenzene	12.40	95	319935	33.85	ug/l	0.00
Spiked Amount 50.000			Recovery	=	67.70%	
Target Compounds						
18) Methyl Acetate	4.33	43	14073	2.10	ug/l	97
20) Methylene Chloride	4.56	84	11105	1.39	ug/l	98
43) Isopropyl Acetate	8.17	43	262949	23.59	ug/l	95


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN090418\
Data File : VN051020.D
Acq On : 4 Sep 2018 15:31
Operator : MD\SY
Sample : PB112553ZHE#06
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 19 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB112553ZHE#06

Quant Time: Sep 05 02:14:48 2018
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N082118W.M
Quant Title : SW846 8260
QLast Update : Wed Aug 22 05:38:12 2018
Response via : Initial Calibration





#1

Pentafluorobenzene

Concen: 50.00 ug/l

RT: 7.67 min Scan# 1891

Delta R.T. 0.00 min

Lab File: VN051020.D

Acq: 4 Sep 2018 15:31

Instrument :

MSVOA_N

ClientSampleId :


PB112553ZHE#06

Tot Ion:168 Resp: 588429

Ion Ratio Lower Upper

168 100

99 50.1 39.6 59.4



Abundance Ion 167.90 (167.60 to 168.60): V

Ion 98.90 (98.60 to 99.60): VN

7.67

250000

200000



150000

100000

50000

0

Time-->



#18

Methyl Acetate

Concen: 2.10 ug/l

RT: 4.33 min Scan# 853

Delta R.T. 0.01 min

Lab File: VN051020.D


Acq: 4 Sep 2018 15:31

Tgt Ion: 43 Resp: 14073

Ion Ratio Lower Upper

43 100

74 23.5 19.8 29.8



Abundance Ion 43.00 (42.70 to 43.70): VN

Ion 74.00 (73.70 to 74.70): VN

4.33

5000

4000


3000


2000

1000

0

Time-->









#20
Methvlene Chloride
Concen: 1.39 ug/l
RT: 4.56 min Scan# 923
Delta R.T. 0.01 min
Lab File: VN051020.D
Acq: 4 Sep 2018 15:31




Instrument :
MSVOA_N
ClientSampleId :
PB112553ZHE#06


Tot Ion: 84 Resp: 11105
Ion Ratio Lower Upper
84 100
49 114.3 93.2 139.8
51 35.3 29.0 43.6
86 65.2 49.8 74.8



#33
1,2-Dichloroethane-d4
Concen: 53.60 ug/l
RT: 8.03 min Scan# 2003
Delta R.T. 0.00 min
Lab File: VN051020.D
Acq: 4 Sep 2018 15:31

Tgt Ion: 65 Resp: 379694
Ion Ratio Lower Upper
65 100
67 54.5 0.0 110.4









#34
 1,4-Difluorobenzene
 Concen: 50.00 ug/l
 RT: 8.59 min Scan# 2177
 Delta R.T. 0.00 min
 Lab File: VN051020.D
 Acq: 4 Sep 2018 15:31

Instrument :
 MSVOA_N
 ClientSampleId :
 PB112553ZHE#06

Tot Ion:114 Resp: 956222
 Ion Ratio Lower Upper
 114 100
 63 19.5 0.0 38.6
 88 15.6 0.0 30.6




Abundance Ion 113.90 (113.60 to 114.60): V
 Ion 63.00 (62.70 to 63.70): VN
 Ion 87.90 (87.60 to 88.60): VN






#35
 Dibromofluoromethane
 Concen: 47.72 ug/l
 RT: 7.59 min Scan# 1867
 Delta R.T. 0.00 min
 Lab File: VN051020.D
 Acq: 4 Sep 2018 15:31

Tgt Ion:113 Resp: 363057
 Ion Ratio Lower Upper
 113 100
 111 102.0 82.6 123.8
 192 23.3 18.0 27.0



Abundance Ion 112.80 (112.50 to 113.50): V
 Ion 110.80 (110.50 to 111.50): V
 Ion 191.70 (191.40 to 192.40): V





#43

Isopropyl Acetate

Concen: 23.59 ug/l

RT: 8.17 min Scan# 2046

Delta R.T. 0.00 min

Lab File: VN051020.D

Acq: 4 Sep 2018 15:31

Instrument :

MSVOA_N

ClientSampleId :

PB112553ZHE#06


Tot Ion: 43 Resp: 262949

Ion Ratio Lower Upper

43 100

61 18.9 17.5 26.3




87 12.6 11.3 16.9



Abundance Ion 43.00 (42.70 to 43.70): VN0

Ion 61.00 (60.70 to 61.70): VN0

Ion 87.00 (86.70 to 87.70): VN0



#50

Toluene-d8

Concen: 45.54 ug/l

RT: 10.09 min Scan# 2645

Delta R.T. 0.00 min

Lab File: VN051020.D


Acq: 4 Sep 2018 15:31

Tgt Ion: 98 Resp: 1295106

Ion Ratio Lower Upper



98 100


100 63.3 51.5 77.3



Abundance Ion 98.00 (97.70 to 98.70): VN0

Ion 100.00 (99.70 to 100.70): VN0









#62
4-Bromofluorobenzene
Concen: 33.85 ug/l
RT: 12.40 min Scan# 3363
Delta R.T. 0.00 min
Lab File: VN051020.D
Acq: 4 Sep 2018 15:31




Instrument :
MSVOA_N
ClientSampleId :
PB112553ZHE#06


Tot Ion	95	Resp	319935
Ion Ratio	Lower	Upper	
95	100		
174	93.7	0.0	183.0
176	89.4	0.0	178.8



#63
Chlorobenzene-d5
Concen: 50.00 ug/l
RT: 11.41 min Scan# 3054
Delta R.T. 0.00 min
Lab File: VN051020.D
Acq: 4 Sep 2018 15:31

Tgt Ion	117	Resp	785933
Ion Ratio	Lower	Upper	
117	100		
82	53.0	42.6	64.0
119	32.8	25.8	38.6





#72

1,4-Dichlorobenzene-d4

Concen: 50.00 ug/l

RT: 13.35 min Scan# 3656

Delta R.T. 0.00 min

Lab File: VN051020.D

Acq: 4 Sep 2018 15:31

Instrument :

MSVOA_N

ClientSampleId :

PB112553ZHE#06


Tot Ion:152 Resp: 231207

Ion Ratio Lower Upper

152 100

115 55.9 27.8 83.4

150 156.8 0.0 351.4



Abundance

Ion 151.90 (151.60 to 152.60): V

Ion 114.90 (114.60 to 115.60): V

Ion 149.90 (149.60 to 150.60): V

