

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN090420\
 Data File : VN063238.D
 Acq On : 4 Sep 2020 11:25
 Operator : JC/MD
 Sample : L3902-03DL 4X
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 SA-MW184I-20200901DL

Quant Time: Sep 08 03:04:59 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N083120W.M
 Quant Title : SW846 8260
 QLast Update : Tue Sep 01 04:21:28 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.62	168	264221	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.55	114	506236	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.38	117	498731	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.32	152	198709	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.98	65	207927	48.14	ug/l	0.00
Spiked Amount 50.000			Recovery	=	96.28%	
35) Dibromofluoromethane	7.55	113	165145	50.85	ug/l	0.00
Spiked Amount 50.000			Recovery	=	101.70%	
50) Toluene-d8	10.06	98	642830	51.42	ug/l	0.00
Spiked Amount 50.000			Recovery	=	102.84%	
62) 4-Bromofluorobenzene	12.38	95	221941	50.75	ug/l	0.00
Spiked Amount 50.000			Recovery	=	101.50%	
Target Compounds						
6) Chloroethane	2.68	64	14379	4.321	ug/l #	88
12) 1,1-Dichloroethene	3.70	96	10680	3.684	ug/l	99
24) 1,1-Dichloroethane	5.80	63	328774	51.692	ug/l	99
32) 1,1,1-Trichloroethane	7.53	97	15437	2.692	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN090420\
Data File : VN063238.D
Acq On : 4 Sep 2020 11:25
Operator : JC/MD
Sample : L3902-03DL 4X
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 6 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
SA-MW184I-20200901DL

Quant Time: Sep 08 03:04:59 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N083120W.M
Quant Title : SW846 8260
QLast Update : Tue Sep 01 04:21:28 2020
Response via : Initial Calibration

