

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN090420\
 Data File : VN063240.D
 Acq On : 4 Sep 2020 12:13
 Operator : JC/MD
 Sample : L3923-03
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 BP-VPB181-EB-20200902

Quant Time: Sep 07 00:43:36 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N083120W.M
 Quant Title : SW846 8260
 QLast Update : Tue Sep 01 04:21:28 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.63	168	260846	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.55	114	508831	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.38	117	496468	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.32	152	194570	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.99	65	207577	48.68	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.36%	
35) Dibromofluoromethane	7.55	113	163711	50.15	ug/l	0.00
Spiked Amount 50.000			Recovery	=	100.30%	
50) Toluene-d8	10.06	98	636883	50.69	ug/l	0.00
Spiked Amount 50.000			Recovery	=	101.38%	
62) 4-Bromofluorobenzene	12.38	95	216243	49.20	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.40%	
Target Compounds						
16) Acetone	3.79	43	10221	8.432	ug/l	99
52) Toluene	10.13	92	24807	2.770	ug/l	93

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN090420\
Data File : VN063240.D
Acq On : 4 Sep 2020 12:13
Operator : JC/MD
Sample : L3923-03
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
BP-VPB181-EB-20200902

Quant Time: Sep 07 00:43:36 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N083120W.M
Quant Title : SW846 8260
QLast Update : Tue Sep 01 04:21:28 2020
Response via : Initial Calibration

