

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN090420\
 Data File : VN063243.D
 Acq On : 4 Sep 2020 13:25
 Operator : JC/MD
 Sample : L3923-06
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 BP-VPB181-GW-898-900

Quant Time: Sep 07 00:52:00 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N083120W.M
 Quant Title : SW846 8260
 QLast Update : Tue Sep 01 04:21:28 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.63	168	268490	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.55	114	526623	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.38	117	521663	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.32	152	200290	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.99	65	215310	49.06	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.12%	
35) Dibromofluoromethane	7.55	113	169294	50.11	ug/l	0.00
Spiked Amount 50.000			Recovery	=	100.22%	
50) Toluene-d8	10.06	98	664942	51.13	ug/l	0.00
Spiked Amount 50.000			Recovery	=	102.26%	
62) 4-Bromofluorobenzene	12.38	95	220780	48.53	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.06%	
Target Compounds						
16) Acetone	3.78	43	13082	10.485	ug/l	Qvalue 97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN090420\
Data File : VN063243.D
Acq On : 4 Sep 2020 13:25
Operator : JC/MD
Sample : L3923-06
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
BP-VPB181-GW-898-900

Quant Time: Sep 07 00:52:00 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N083120W.M
Quant Title : SW846 8260
QLast Update : Tue Sep 01 04:21:28 2020
Response via : Initial Calibration

