

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN090420\
 Data File : VN063244.D
 Acq On : 4 Sep 2020 13:49
 Operator : JC/MD
 Sample : L3902-14
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 SA-MW132S-20200901

Quant Time: Sep 07 00:55:12 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N083120W.M
 Quant Title : SW846 8260
 QLast Update : Tue Sep 01 04:21:28 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.63	168	263903	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.55	114	521752	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.38	117	510298	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.32	152	197557	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.99	65	215384	49.93	ug/l	0.00
Spiked Amount 50.000			Recovery	=	99.86%	
35) Dibromofluoromethane	7.55	113	168109	50.22	ug/l	0.00
Spiked Amount 50.000			Recovery	=	100.44%	
50) Toluene-d8	10.06	98	656956	50.99	ug/l	0.00
Spiked Amount 50.000			Recovery	=	101.98%	
62) 4-Bromofluorobenzene	12.38	95	225116	49.95	ug/l	0.00
Spiked Amount 50.000			Recovery	=	99.90%	
Target Compounds						
24) 1,1-Dichloroethane	5.80	63	7808	1.229	ug/l #	Qvalue 94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN090420\
Data File : VN063244.D
Acq On : 4 Sep 2020 13:49
Operator : JC/MD
Sample : L3902-14
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
SA-MW132S-20200901

Quant Time: Sep 07 00:55:12 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N083120W.M
Quant Title : SW846 8260
QLast Update : Tue Sep 01 04:21:28 2020
Response via : Initial Calibration

