

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN090420\
 Data File : VN063250.D
 Acq On : 4 Sep 2020 16:13
 Operator : JC/MD
 Sample : PB131437TB
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB131437TB

Manual Integrations
 APPROVED

MMDadoda
 9/8/2020 1:33:38 PM

Quant Time: Sep 07 01:34:22 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N083120W.M
 Quant Title : SW846 8260
 QLast Update : Tue Sep 01 04:21:28 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.63	168	274972	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.55	114	532878	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.38	117	536541	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.32	152	213614	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.99	65	217151	48.31	ug/l	0.00
Spiked Amount 50.000			Recovery	=	96.62%	
35) Dibromofluoromethane	7.55	113	167650	49.04	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.08%	
50) Toluene-d8	10.06	98	692041	52.59	ug/l	0.00
Spiked Amount 50.000			Recovery	=	105.18%	
62) 4-Bromofluorobenzene	12.38	95	238670	51.85	ug/l	0.00
Spiked Amount 50.000			Recovery	=	103.70%	
Target Compounds						
16) Acetone	3.78	43	19817	15.509	ug/l	98
18) Methyl Acetate	4.29	43	3778m	1.153	ug/l	
20) Methylene Chloride	4.51	84	11206	2.653	ug/l	87
43) Isopropyl Acetate	8.13	43	92588	11.082	ug/l #	92

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN090420\
Data File : VN063250.D
Acq On : 4 Sep 2020 16:13
Operator : JC/MD
Sample : PB131437TB
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 18 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB131437TB

Manual Integrations
APPROVED

MMDadoda
9/8/2020 1:33:38 PM

Quant Time: Sep 07 01:34:22 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N083120W.M
Quant Title : SW846 8260
QLast Update : Tue Sep 01 04:21:28 2020
Response via : Initial Calibration

