

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN090420\
 Data File : VN063251.D
 Acq On : 4 Sep 2020 16:37
 Operator : JC/MD
 Sample : L3871-05
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 TR-05-090220

Quant Time: Sep 07 01:36:50 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N083120W.M
 Quant Title : SW846 8260
 QLast Update : Tue Sep 01 04:21:28 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.63	168	260314	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.55	114	519395	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.38	117	508268	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.32	152	196118	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.99	65	212225	49.87	ug/l	0.00
Spiked Amount 50.000			Recovery	=	99.74%	
35) Dibromofluoromethane	7.55	113	159129	47.76	ug/l	0.00
Spiked Amount 50.000			Recovery	=	95.52%	
50) Toluene-d8	10.06	98	649502	50.64	ug/l	0.00
Spiked Amount 50.000			Recovery	=	101.28%	
62) 4-Bromofluorobenzene	12.38	95	218997	48.81	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.62%	
Target Compounds						
16) Acetone	3.78	43	11224	9.278	ug/l	93
20) Methylene Chloride	4.51	84	8756	2.103	ug/l	94
95) Naphthalene	15.11	128	57799	5.113	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN090420\
Data File : VN063251.D
Acq On : 4 Sep 2020 16:37
Operator : JC/MD
Sample : L3871-05
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 19 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
TR-05-090220

Quant Time: Sep 07 01:36:50 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N083120W.M
Quant Title : SW846 8260
QLast Update : Tue Sep 01 04:21:28 2020
Response via : Initial Calibration

