

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN090420\
 Data File : VN063266.D
 Acq On : 5 Sep 2020 4:00
 Operator : JC/MD
 Sample : L3877-08
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 38 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 TP-2

Manual Integrations
 APPROVED

MMDadoda
 9/8/2020 1:32:34 PM

Quant Time: Sep 07 02:36:44 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N083120W.M
 Quant Title : SW846 8260
 QLast Update : Tue Sep 01 04:21:28 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.63	168	291817	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.55	114	554245	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.38	117	546848	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.32	152	221739	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.99	65	227506	47.69	ug/l	0.00
Spiked Amount 50.000			Recovery	=	95.38%	
35) Dibromofluoromethane	7.55	113	172258	48.44	ug/l	0.00
Spiked Amount 50.000			Recovery	=	96.88%	
50) Toluene-d8	10.06	98	700529	51.19	ug/l	0.00
Spiked Amount 50.000			Recovery	=	102.38%	
62) 4-Bromofluorobenzene	12.38	95	252607	52.76	ug/l	0.00
Spiked Amount 50.000			Recovery	=	105.52%	
Target Compounds						
16) Acetone	3.78	43	30329	22.365	ug/l	95
18) Methyl Acetate	4.29	43	4930m	1.418	ug/l	
20) Methylene Chloride	4.51	84	19691	4.719	ug/l	96
40) Benzene	8.00	78	117080	7.477	ug/l	98
43) Isopropyl Acetate	8.13	43	83160	9.570	ug/l #	89
95) Naphthalene	15.10	128	486302	38.049	ug/l	100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN090420\
Data File : VN063266.D
Acq On : 5 Sep 2020 4:00
Operator : JC/MD
Sample : L3877-08
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 38 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
TP-2

Manual Integrations
APPROVED

MMDadoda
9/8/2020 1:32:34 PM

Quant Time: Sep 07 02:36:44 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N083120W.M
Quant Title : SW846 8260
QLast Update : Tue Sep 01 04:21:28 2020
Response via : Initial Calibration

