

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN090424\
 Data File : VN083649.D
 Acq On : 04 Sep 2024 15:06
 Operator : JC\MD
 Sample : P3777-05RE
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 MW-04-12-082824RE

Quant Time: Sep 05 02:26:45 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N083024W.M
 Quant Title : SW846 8260
 QLast Update : Sat Aug 31 00:48:10 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.224	168	144230	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.100	114	296531	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	321093	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.788	152	157118	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.577	65	112725	55.922	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	111.840%
35) Dibromofluoromethane	8.165	113	92968	52.090	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	104.180%
50) Toluene-d8	10.565	98	333551	47.985	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	95.960%
62) 4-Bromofluorobenzene	12.847	95	158486	61.556	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	=	123.120%#

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN090424\
Data File : VN083649.D
Acq On : 04 Sep 2024 15:06
Operator : JC\MD
Sample : P3777-05RE
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
MW-04-12-082824RE

Quant Time: Sep 05 02:26:45 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N083024W.M
Quant Title : SW846 8260
QLast Update : Sat Aug 31 00:48:10 2024
Response via : Initial Calibration

