

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN090425\
 Data File : VN087719.D
 Acq On : 04 Sep 2025 13:56
 Operator : JC\MD
 Sample : Q3017-02
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 TT-076-IDWGW-20250903

Quant Time: Sep 05 02:20:41 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\624N080525W.M
 Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
 QLast Update : Wed Aug 06 02:01:21 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Bromochloromethane	7.794	128	31627	30.000	ug/l	0.00
28) 1,4-Difluorobenzene	9.082	114	160432	30.000	ug/l	0.00
57) Chlorobenzene-d5	11.847	117	142263	30.000	ug/l	0.00
System Monitoring Compounds						
27) 1,2-Dichloroethane-d4	8.559	65	86846	29.771	ug/l	0.00
Spiked Amount	30.000	Range	91 - 110	Recovery	=	99.233%
60) 4-Bromofluorobenzene	12.829	95	62220	25.414	ug/l	0.00
Spiked Amount	30.000	Range	63 - 112	Recovery	=	84.700%
63) Toluene-d8	10.547	98	193077	29.496	ug/l	0.00
Spiked Amount	30.000	Range	91 - 112	Recovery	=	98.333%
Target Compounds						
37) Trichloroethene	9.335	130	12183	6.696	ug/l	Qvalue 99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN090425\
Data File : VN087719.D
Acq On : 04 Sep 2025 13:56
Operator : JC\MD
Sample : Q3017-02
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
TT-076-IDWGW-20250903

Quant Time: Sep 05 02:20:41 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\624N080525W.M
Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
QLast Update : Wed Aug 06 02:01:21 2025
Response via : Initial Calibration

