

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN090425\
 Data File : VN087724.D
 Acq On : 04 Sep 2025 15:41
 Operator : JC\MD
 Sample : BLK
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 BLK

Quant Time: Sep 05 02:22:55 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\624N080525W.M
 Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
 QLast Update : Wed Aug 06 02:01:21 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Bromochloromethane	7.800	128	39657	30.000	ug/l	0.00
28) 1,4-Difluorobenzene	9.083	114	194323	30.000	ug/l	0.00
57) Chlorobenzene-d5	11.847	117	169955	30.000	ug/l	0.00

System Monitoring Compounds						
27) 1,2-Dichloroethane-d4	8.565	65	103885	28.401	ug/l	0.00
Spiked Amount	30.000	Range	91 - 110	Recovery	=	94.667%
60) 4-Bromofluorobenzene	12.829	95	74876	25.600	ug/l	0.00
Spiked Amount	30.000	Range	63 - 112	Recovery	=	85.333%
63) Toluene-d8	10.547	98	232470	29.727	ug/l	0.00
Spiked Amount	30.000	Range	91 - 112	Recovery	=	99.100%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN090425\
Data File : VN087724.D
Acq On : 04 Sep 2025 15:41
Operator : JC\MD
Sample : BLK
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
BLK

Quant Time: Sep 05 02:22:55 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\624N080525W.M
Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
QLast Update : Wed Aug 06 02:01:21 2025
Response via : Initial Calibration

