

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN090519\
Data File : VN057695.D
Acq On : 5 Sep 2019 10:53
Operator : JC/SP
Sample : VN0905WBL01
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VN0905WBL01

Quant Time: Sep 06 01:35:20 2019
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA N\METHODS\624N090319W.M
Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
QLast Update : Wed Sep 04 02:08:50 2019
Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Bromochloromethane	7.18	128	58422	30.00	ug/l	0.00
28) 1,4-Difluorobenzene	8.58	114	340304	30.00	ug/l	0.00
57) Chlorobenzene-d5	11.41	117	241324	30.00	ug/l	0.00

System Monitoring Compounds

27) 1,2-Dichloroethane-d4	8.01	65	187028	30.11	ug/l	0.00
Spiked Amount	30.000	Range	50 - 169	Recovery	=	100.37%
60) 4-Bromofluorobenzene	12.40	95	77170	18.72	ug/l	0.00
Spiked Amount	30.000	Range	56 - 143	Recovery	=	62.40%
63) Toluene-d8	10.09	98	410615	35.34	ug/l	0.00
Spiked Amount	30.000	Range	66 - 137	Recovery	=	117.80%

Target Compounds	Ovalue
------------------	--------

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN090519\
Data File : VN057695.D
Acq On : 5 Sep 2019 10:53
Operator : JC/SP
Sample : VN0905WBL01
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VN0905WBL01

Quant Time: Sep 06 01:35:20 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA N\METHODS\624N090319W.M
Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
QLast Update : Wed Sep 04 02:08:50 2019
Response via : Initial Calibration

