

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN090524\
 Data File : VN083684.D
 Acq On : 05 Sep 2024 18:39
 Operator : JC\MD
 Sample : P3802-08RE
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 OWBR-02-170-082924RE

Quant Time: Sep 06 02:21:47 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N083024W.M
 Quant Title : SW846 8260
 QLast Update : Sat Aug 31 00:48:10 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.224	168	134180	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.100	114	276843	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	303945	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	146151	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.577	65	107210	57.170	ug/l	0.00
Spiked Amount	50.000	Range 74 - 125	Recovery	=	114.340%	
35) Dibromofluoromethane	8.165	113	85949	51.582	ug/l	0.00
Spiked Amount	50.000	Range 75 - 124	Recovery	=	103.160%	
50) Toluene-d8	10.565	98	306409	47.233	ug/l	0.00
Spiked Amount	50.000	Range 86 - 113	Recovery	=	94.460%	
62) 4-Bromofluorobenzene	12.847	95	146857	61.096	ug/l	0.00
Spiked Amount	50.000	Range 77 - 121	Recovery	=	122.200%#	
Target Compounds						
					Qvalue	
27) cis-1,2-Dichloroethene	7.488	96	1898	1.039	ug/l	86
44) Trichloroethene	9.347	130	6830	3.632	ug/l	98
51) 4-Methyl-2-Pentanone	10.447	43	1184	0.503	ug/l	93
52) Toluene	10.629	92	3311	0.677	ug/l	84
69) o-Xylene	12.400	106	3877	0.875	ug/l	100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN090524\
Data File : VN083684.D
Acq On : 05 Sep 2024 18:39
Operator : JC\MD
Sample : P3802-08RE
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 22 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
OWBR-02-170-082924RE

Quant Time: Sep 06 02:21:47 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N083024W.M
Quant Title : SW846 8260
QLast Update : Sat Aug 31 00:48:10 2024
Response via : Initial Calibration

